



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108138

Slope : 98.7

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2628         |
| buffer solution pH 7 yellow    | W2626         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2632         |
| buffer solution pH 7 yellow    | W2569         |
| Buffer Solution, PH2 (500ml)   | W2592         |
| Buffer Solution, PH12 (500ml)  | W2616         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.2             | 4.01        | 03/05/2020 | 16:00     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.2             | 7.01        | 03/05/2020 | 16:01     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.2             | 10.03       | 03/05/2020 | 16:02     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.2             | 7.01        | 03/05/2020 | 16:03     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.2             | 2.01        | 03/05/2020 | 17:15     |
| 6   | L1803-01    | 1  | Water  | NA          | NA          | 13.9             | 7.30        | 03/05/2020 | 17:16     |
| 7   | L1803-01DUP | 1  | Water  | NA          | NA          | 13.8             | 7.31        | 03/05/2020 | 17:17     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.1             | 12.02       | 03/05/2020 | 17:18     |

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1806

WorkList ID : 136764

Department : Wet-Chemistry

Date : 03-05-2020 17:12:04

| Due Date   | Matrix | Sample   | Test                                   | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|----------------------------------------|--------------|----------|------------------|-----------------|--------------|--------|
| 03/05/2020 | Water  | L1803-01 | <input checked="" type="checkbox"/> pH | Cool 4 deg C | PSEG01   | L41              | N38129          | 03/05/2020   | 9040C  |

Date/Time

03-05-2020 17:15

Received by:

J0

Relinquished by:

J0

Date/Time

03-05-2020 18:15

Received by:

J0

Relinquished by:

J0