



Analytical Summary Report

Analysis Method: 9050A

ANALYST: Alexander

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB108141

Meter ID: WC Conductance meter-1

Thermometer ID: Ph meter-Temp probe-1

Reagent/Standard	Lot/Log #
conductivity standard	WP81395

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB108141BL		1	22.1	0.80	0.85	03/05/2020	15:20
2	LB108141BS	1413	1	21.7	1312.00	1400.26	03/05/2020	15:22
3	L1801-09		1	21.8	777.00	827.58	03/05/2020	15:26

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$