



## Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB108155

| Reagent/Standard                 | Lot/Log # |
|----------------------------------|-----------|
| TURBIDITY CCV 10 NTU             | WP82759   |
| Turbidity Calibration std, 0NTU  | WP82760   |
| Turbidity Calibration std, 20NTU | WP82755   |
| Turbidity Calibration std, 40NTU | WP82754   |
| Turbidity Calibration std, 1NTU  | WP82758   |
| Turbidity Calibration std, 10NTU | WP82756   |
| Turbidity Calibration std, 5NTU  | WP82757   |
| 10 NTU Standard 500 ml           | W2489     |

Intercept: 0.0512

Slope: 0.843

Regression: 1

| Seq | Lab ID | True Value (NTU) | Dilution | Reading | Anal Date  | Anal Time |
|-----|--------|------------------|----------|---------|------------|-----------|
| 1   | CAL1   | 0                | 1        | 0.152   | 03/06/2020 | 13:22     |
| 2   | CAL2   | 1                | 1        | 0.934   | 03/06/2020 | 13:25     |
| 3   | CAL3   | 5                | 1        | 4.140   | 03/06/2020 | 13:27     |
| 4   | CAL4   | 10               | 1        | 8.450   | 03/06/2020 | 13:28     |
| 5   | CAL5   | 20               | 1        | 16.900  | 03/06/2020 | 13:29     |
| 6   | CAL6   | 40               | 1        | 33.800  | 03/06/2020 | 13:33     |



## Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB108155

| Seq | Lab ID     | True Value<br>(NTU) | Dilution | Reading | Result<br>(NTU) | AnalDate   | Anal Time |
|-----|------------|---------------------|----------|---------|-----------------|------------|-----------|
| 1   | ICV        | 10                  | 1        | 9.100   | 10.734          | 03/06/2020 | 13:36     |
| 2   | ICB        |                     | 1        | 0.150   | 0.117           | 03/06/2020 | 13:38     |
| 3   | CCV1       | 10                  | 1        | 8.480   | 9.999           | 03/06/2020 | 13:40     |
| 4   | CCB1       |                     | 1        | 0.160   | 0.129           | 03/06/2020 | 13:43     |
| 5   | LB108155BL |                     | 1        | 0.155   | 0.123           | 03/06/2020 | 13:45     |
| 6   | L1801-37   |                     | 1        | 2.980   | 3.474           | 03/06/2020 | 13:50     |
| 7   | CCV2       | 10                  | 1        | 8.440   | 9.951           | 03/06/2020 | 13:55     |
| 8   | CCB2       |                     | 1        | 0.149   | 0.116           | 03/06/2020 | 13:58     |