



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB108156

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP82759
Turbidity Calibration std, 0NTU	WP82760
Turbidity Calibration std, 20NTU	WP82755
Turbidity Calibration std, 40NTU	WP82754
Turbidity Calibration std, 1NTU	WP82758
Turbidity Calibration std, 10NTU	WP82756
Turbidity Calibration std, 5NTU	WP82757
10 NTU Standard 500 ml	W2489

Intercept: 0.0512

Slope: 0.843

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.152	03/06/2020	13:22
2	CAL2	1	1	0.934	03/06/2020	13:25
3	CAL3	5	1	4.140	03/06/2020	13:27
4	CAL4	10	1	8.450	03/06/2020	13:28
5	CAL5	20	1	16.900	03/06/2020	13:29
6	CAL6	40	1	33.800	03/06/2020	13:33



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB108156

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.100	10.734	03/06/2020	13:36
2	ICB		1	0.150	0.117	03/06/2020	13:38
3	CCV1	10	1	8.440	9.951	03/06/2020	13:55
4	CCB1		1	0.149	0.116	03/06/2020	13:58
5	LB108156BL		1	0.153	0.121	03/06/2020	14:02
6	L1801-38		1	2.970	3.462	03/06/2020	14:08
7	CCV2	10	1	8.400	9.904	03/06/2020	14:11
8	CCB2		1	0.150	0.117	03/06/2020	14:13