



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108158

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.00	03/06/2020	14:35
2	CAL2	1	Water	NA	NA	20.1	7.00	03/06/2020	14:36
3	CAL3	1	Water	NA	NA	20.1	10.03	03/06/2020	14:37
4	ICV	1	Water	NA	NA	20.2	6.99	03/06/2020	14:39
5	CCV1	1	Water	NA	NA	20.2	2.01	03/06/2020	14:40
6	L1837-01	1	Solid	20.03	20	20.7	6.12	03/06/2020	14:42
7	L1837-01DUP	1	Solid	20.04	20	20.8	6.13	03/06/2020	14:44
8	CCV2	1	Water	NA	NA	20.2	12.02	03/06/2020	14:45

183108158

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1737 WorkList ID : 136807 Department : Wet-Chemistry Date : 03-06-2020 14:26:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/06/2020	Solid	L1837-01	pH	Cool 4 deg C	TULL02	O61	FK-PS-01-065	03/06/2020	9045D

Date/Time 03-06-2020 14:30
Received by: JS
Relinquished by: JS

Date/Time 03-06-2020 15:30
Received by: JS
Relinquished by: JS