



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB108159

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	03/06/2020	13:40
2	L1803-02		1	8	100.00	>212.0	>212.0	03/06/2020	14:10
3	L1803-02DUP		1	8	100.00	>212.0	>212.0	03/06/2020	14:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB108159

WorkList Name : fp3620

WorkList ID : 136802

Department : Wet-Chemistry

Date : 03-06-2020 13:10:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/12/2020	Water	L1803-02	C	Cool 4 deg C	PSEG01	L41	N38129	03/05/2020	1010A

Date/Time 03/06/2020 13:15
 Received by: KS
 Relinquished by: fe

Date/Time 03/06/2020 14:40
 Received by: fe
 Relinquished by: KS