


LB 108180

ISM02.4

INSTRUMENT ID: CV1

 μ Abs.:

42698



Linear

A= 0.0000e+000

B= 2.3579e-004

C= 1.1173e-002

Rho= 0.9997956

Accept=Accepted

Slope
Intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.021	0.021	40	0.000	40					—
0.05	0.050					149					—
0.20	0.200	0.176	-0.024	698	0.0 %	698					-12
2.50	2.500	2.601	0.101	10985	0.0 %	10985					4
5.00	5.000	4.899	-0.101	20728	0.0 %	20728					-2
7.50	7.500	7.425	-0.075	31443	0.0 %	31443					-1
10.0	10.000	10.079	0.079	42698	0.0 %	42698					1

LB108180
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	40	-		0 ISM02.4	PPB	3/9/2020 16:43	Std
	0.2 S0.2	698	-		0.2 ISM02.4	PPB	3/9/2020 16:46	Std
	2.5 S2.5	10985	-		2.5 ISM02.4	PPB	3/9/2020 16:48	Std
	5 S5.0	20728	-		5 ISM02.4	PPB	3/9/2020 16:50	Std
	7.5 S7.5	31443	-		7.5 ISM02.4	PPB	3/9/2020 16:52	Std
	10 S10.0	42698	-		10 ISM02.4	PPB	3/9/2020 16:55	Std
ICV	ICV	15983	3.7797 -		ISM02.4	PPB	3/9/2020 17:03	SMPL
ICB	ICB	-81	-0.0079 -		ISM02.4	PPB	3/9/2020 17:05	SMPL
CCV085	CCV085	21106	4.9877 -		ISM02.4	PPB	3/9/2020 17:07	SMPL
CCB085	CCB085	-127	-0.0188 -		ISM02.4	PPB	3/9/2020 17:09	SMPL
PB127382BL	PBW382	39	0.0204 -		ISM02.4	PPB	3/9/2020 17:12	SMPL
L1672-10	MBFM25	49	0.0227 -		ISM02.4	PPB	3/9/2020 17:14	SMPL
L1787-04	MCOAB6	67	0.027 -		ISM02.4	PPB	3/9/2020 17:16	SMPL
CCV086	CCV086	20826	4.9217 -		ISM02.4	PPB	3/9/2020 17:18	SMPL
CCB086	CCB086	-101	-0.0126 -		ISM02.4	PPB	3/9/2020 17:21	SMPL