



Analytical Summary Report

Analysis Method: 9045D
Parameter: Corrosivity
Run Number: LB108181
BalanceID: WC SC-4

Analyst By : jignesh
Supervisor Review By : apatel
Slope : 98.5
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/09/2020	16:40
2	CAL2	1	Water	NA	NA	20.2	7.01	03/09/2020	16:41
3	CAL3	1	Water	NA	NA	20.2	10.03	03/09/2020	16:42
4	ICV	1	Water	NA	NA	20.1	7.00	03/09/2020	16:45
5	CCV1	1	Water	NA	NA	20.1	2.01	03/09/2020	16:46
6	L1849-02	1	Solid	20.03	20	21.0	8.50	03/09/2020	16:47
7	L1849-04	1	Solid	20.04	20	21.3	8.34	03/09/2020	16:49
8	L1852-04	1	Solid	20.01	20	21.4	8.75	03/09/2020	16:50
9	L1852-08	1	Solid	20.03	20	21.2	8.37	03/09/2020	17:00
10	L1852-12	1	Solid	20.04	20	20.9	7.96	03/09/2020	17:02
11	L1852-16	1	Solid	20.03	20	21.2	7.58	03/09/2020	17:05
12	L1852-20	1	Solid	20.04	20	20.9	8.78	03/09/2020	17:06
13	L1852-24	1	Solid	20.03	20	21.4	9.08	03/09/2020	17:09
14	L1852-24DUP	1	Solid	20.04	20	21.5	9.09	03/09/2020	17:10
15	CCV2	1	Water	NA	NA	20.1	12.02	03/09/2020	17:12

YB 108181

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1852 WorkList ID : 136843 Department : Wet-Chemistry Date : 03-09-2020 15:36:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/09/2020	Solid	L1849-02	Corrosivity	Cool 4 deg C	PSEG01	I52	VNJ-230-218-211-216-242	03/09/2020	9045D
03/09/2020	Solid	L1849-04	Corrosivity	Cool 4 deg C	PSEG01	I52	VNJ-215-72-12018	03/09/2020	9045D
03/07/2020	Solid	L1852-04	Corrosivity	Cool 4 deg C	PSEG01	I52	TP-1-S	03/07/2020	9045D
03/07/2020	Solid	L1852-08	Corrosivity	Cool 4 deg C	PSEG01	I52	TP-2-S	03/07/2020	9045D
03/07/2020	Solid	L1852-12	Corrosivity	Cool 4 deg C	PSEG01	I52	TP-2-D	03/07/2020	9045D
03/07/2020	Solid	L1852-16	Corrosivity	Cool 4 deg C	PSEG01	I52	TP-3-S	03/07/2020	9045D
03/07/2020	Solid	L1852-20	Corrosivity	Cool 4 deg C	PSEG01	I52	TP-3-D	03/07/2020	9045D
03/07/2020	Solid	L1852-24	Corrosivity	Cool 4 deg C	PSEG01	I52	TP-4-S	03/07/2020	9045D

Date/Time 03-09-2020 15:35
 Received by: JO
 Relinquished by: JC

Date/Time 03-09-2020 17:10
 Received by: JC
 Relinquished by: JP