



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108183

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/09/2020	16:40
2	CAL2	1	Water	NA	NA	20.2	7.01	03/09/2020	16:41
3	CAL3	1	Water	NA	NA	20.2	10.04	03/09/2020	16:42
4	ICV	1	Water	NA	NA	20.1	7.00	03/09/2020	16:45
5	CCV1	1	Water	NA	NA	20.2	2.01	03/09/2020	16:46
6	L1801-21	1	Water	NA	NA	21.7	8.02	03/09/2020	17:15
7	L1852-25	1	Water	NA	NA	15.4	6.46	03/09/2020	17:15
8	L1852-25DUP	1	Water	NA	NA	15.5	6.47	03/09/2020	17:16
9	CCV2	1	Water	NA	NA	20.2	12.02	03/09/2020	17:17

VB108183

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1801 WorkList ID : 136846 Department : Wet-Chemistry Date : 03-09-2020 15:38:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/07/2020	Water	L1852-25	pH	Cool 4 deg C	PSEG01	I52	TP-2	03/07/2020	9040C

Date/Time 03-09-2020 15:19
Received by: AC
Relinquished by: AC

Date/Time 03-09-2020 17:10
Received by: AC
Relinquished by: AC