



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108185

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/09/2020	16:40
2	CAL2	1	Water	NA	NA	20.2	7.01	03/09/2020	16:41
3	CAL3	1	Water	NA	NA	20.2	10.04	03/09/2020	16:42
4	ICV	1	Water	NA	NA	20.1	7.00	03/09/2020	16:45
5	CCV1	1	Water	NA	NA	20.2	2.01	03/09/2020	16:46
6	L1801-20	1	Water	NA	NA	21.7	8.01	03/09/2020	16:49
7	CCV2	1	Water	NA	NA	20.1	12.02	03/09/2020	17:00