



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB108195

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	03/10/2020	09:50
2	L1852-26		1	8	100.00	>212.0	>212.0	03/10/2020	10:20
3	L1852-26DUP		1	8	100.00	>212.0	>212.0	03/10/2020	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 10 8195

WorkList Name : fp31020

WorkList ID : 136880

Department : Wet-Chemistry

Date : 03-10-2020 09:21:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/16/2020	Water	L1852-26	C Flash Point	Cool 4 deg C	PSEG01	I52	TP-2	03/07/2020	1010A

Date/Time 03/10/2020 09:25
 Received by: KS
 Relinquished by: JG

Date/Time 03/10/2020 10:50
 Received by: JG
 Relinquished by: KS