

LB108196


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53227



A= 0.0000e+000

B= 1.8794e-004

C= -1.2680e-002

Rho= 0.9999948

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	94	0.000	94				
0.2	0.200	0.190	-0.010	1080	0.0 %	1080				
2.5	2.500	2.490	-0.010	13319	0.0 %	13319				
5.0	5.000	5.023	0.023	26797	0.0 %	26797				
7.5	7.500	7.500	0.000	39976	0.0 %	39976				
10.0	10.000	9.991	-0.009	53227	0.0 %	53227				

LB108196 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 10 Mar 2020 12:37:48				
Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type
0.0 - 1	50	94	-	0.0000	7471B	PPB	10 Mar 2020 13:56:17	S Std
0.2 - 1	50.2	1080	-	0.2000	7471B	PPB	10 Mar 2020 13:58:34	S Std
2.5 - 1	52.5	13319	-	2.5000	7471B	PPB	10 Mar 2020 14:00:51	S Std
5.0 - 1	55	26797	-	5.0000	7471B	PPB	10 Mar 2020 14:03:08	S Std
7.5 - 1	57.5	39976	-	7.5000	7471B	PPB	10 Mar 2020 14:05:24	S Std
10.0 - 1	511	53227	-	10.0000	7471B	PPB	10 Mar 2020 14:10:19	S Std
ICV52 - 1	ICV52	21272	3.9851	-	7471B	PPB	10 Mar 2020 14:13:02	U SMPL
ICB52 - 1	ICB52	-190	-0.0484	-	7471B	PPB	10 Mar 2020 14:15:18	U SMPL
CCV92 - 1	CCV92	26652	4.9962	-	7471B	PPB	10 Mar 2020 14:17:34	U SMPL
CCB92 - 1	CCB92	-205	-0.0512	-	7471B	PPB	10 Mar 2020 14:19:49	U SMPL
CRA - 1	CRA	1035	0.1818	-	7471B	PPB	10 Mar 2020 14:22:06	U SMPL
HighStd - 1	HighStd	53277	10.0000	-	7471B	PPB	10 Mar 2020 14:24:20	U SMPL
ChkStd - 1	ChkStd	37569	7.0479	-	7471B	PPB	10 Mar 2020 14:26:36	U SMPL
PB127355BL - 1	PBS	-53	-0.0226	-	7471B	PPB	10 Mar 2020 14:31:28	U SMPL
PB127355BS - 1	LCSS	18863	3.5324	-	7471B	PPB	10 Mar 2020 14:33:44	U SMPL
L1836-01 - 1	VNJ-212-VNJ-221	21990	4.1200	-	7471B	PPB	10 Mar 2020 14:40:55	U SMPL
L1836-01DUP - 1	VNJ-212-VNJ-221DUP	21005	3.9349	-	7471B	PPB	10 Mar 2020 14:45:14	U SMPL
L1836-01MS - 1	VNJ-212-VNJ-221MS	39585	7.4268	-	7471B	PPB	10 Mar 2020 14:47:31	U SMPL
L1836-01MSD - 1	VNJ-212-VNJ-221MSD	39961	7.4974	-	7471B	PPB	10 Mar 2020 14:49:49	U SMPL
L1837-01 - 1	FK-PS-01-065	2820	0.5173	-	7471B	PPB	10 Mar 2020 14:52:06	U SMPL
CCV93 - 1	CCV93	27519	5.1591	-	7471B	PPB	10 Mar 2020 14:54:22	U SMPL
CCB93 - 1	CCB93	-64	-0.0247	-	7471B	PPB	10 Mar 2020 14:56:38	U SMPL
L1844-01 - 1	GP-14-030520	3463	0.6381	-	7471B	PPB	10 Mar 2020 14:58:54	U SMPL
L1844-02 - 1	GP-15-030520	4071	0.7524	-	7471B	PPB	10 Mar 2020 15:01:09	U SMPL
L1844-03 - 1	GP-16-030520	1745	0.3153	-	7471B	PPB	10 Mar 2020 15:03:25	U SMPL
L1844-04 - 1	GP-17-030520	975	0.1706	-	7471B	PPB	10 Mar 2020 15:05:40	U SMPL
L1844-05 - 1	GP-18-030520	1468	0.2632	-	7471B	PPB	10 Mar 2020 15:07:56	U SMPL
L1844-06 - 1	GP-19-030520	3314	0.6101	-	7471B	PPB	10 Mar 2020 15:10:12	U SMPL
L1844-07 - 1	GP-20-030520	6532	1.2149	-	7471B	PPB	10 Mar 2020 15:12:28	U SMPL
L1849-01 - 1	VNJ-230-218-211-216-242	5235	0.9712	-	7471B	PPB	10 Mar 2020 15:14:45	U SMPL
L1849-03 - 1	VNJ-215-72-12018	20211	3.7857	-	7471B	PPB	10 Mar 2020 15:17:01	U SMPL
L1852-03 - 1	TP-1-S	7492	1.3953	-	7471B	PPB	10 Mar 2020 15:19:17	U SMPL
CCV94 - 1	CCV94	27819	5.2155	-	7471B	PPB	10 Mar 2020 15:21:32	U SMPL
CCB94 - 1	CCB94	-33	-0.0189	-	7471B	PPB	10 Mar 2020 15:23:48	U SMPL
L1852-07 - 1	TP-2-S	43930	8.2433	-	7471B	PPB	10 Mar 2020 15:26:03	U SMPL
L1852-11 - 1	TP-2-D	63438	11.9096	-	7471B	PPB	10 Mar 2020 15:28:19	U SMPL
L1852-15 - 1	TP-3-S	7765	1.4466	-	7471B	PPB	10 Mar 2020 15:30:36	U SMPL
L1852-19 - 1	TP-3-D	7154	1.3318	-	7471B	PPB	10 Mar 2020 15:32:53	U SMPL
L1852-23 - 1	TP-4-S	1727	0.3119	-	7471B	PPB	10 Mar 2020 15:35:08	U SMPL
L1836-01LX5 - 1		4358	0.8063	-	7471B	PPB	10 Mar 2020 15:37:24	U SMPL
L1836-01A - 1		45580	8.5534	-	7471B	PPB	10 Mar 2020 15:39:41	U SMPL
L1852-11DLX2 - 1		32182	6.0355	-	7471B	PPB	10 Mar 2020 15:41:58	U SMPL
CCV95 - 1	CCV95	28004	5.2503	-	7471B	PPB	10 Mar 2020 15:44:16	U SMPL
CCB95 - 1	CCB95	-44	-0.0209	-	7471B	PPB	10 Mar 2020 15:46:33	U SMPL