

Analysis Method: SM4500 S E

ANALYST: ketankumar

Parameter: Sulfide

SUPERVISOR REVIEW BY: apatel

Run Number: LB108203

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP81413
IODINE SOLUTION .025N 1L	W2669
Starch Solution, 4L	W2379

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	T1 (mL)	T2 Initial	T2 Final	T2 Diff. (mL)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	AnalDate	Anal Time
1	PB127424BL		1	50	50	5.00	0.00	4.90	4.90	0.10	0.00	0.00	03/10/2020	16:25
2	PB127424BS	25	1	50	50	5.00	0.00	2.28	2.28	2.72	2.62	20.96	03/10/2020	16:28
3	L1801-32		1	50	50	5.00	0.00	4.00	4.00	1.00	0.90	7.20	03/10/2020	16:32

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume