


LB108205

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	123	0.000	123				
0.2	0.200	0.207	0.007	981	0.0 %	981				
2.5	2.500	2.423	-0.077	11523	0.0 %	11523				
5.0	5.000	5.059	0.059	24060	0.0 %	24060				
7.5	7.500	7.484	-0.016	35597	0.0 %	35597				
10.0	10.000	10.001	0.001	47571	0.0 %	47571				

LB108205 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Mar 2020 08:46:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	123	-	0.0000	7471B	PPB	11 Mar 2020 08:49:36	S	Std
0.2 - 1	50.2	981	-	0.2000	7471B	PPB	11 Mar 2020 08:51:53	S	Std
2.5 - 1	52.5	11523	-	2.5000	7471B	PPB	11 Mar 2020 08:54:09	S	Std
5.0 - 1	55	24060	-	5.0000	7471B	PPB	11 Mar 2020 08:56:26	S	Std
7.5 - 1	57.5	35597	-	7.5000	7471B	PPB	11 Mar 2020 08:58:43	S	Std
10.0 - 1	510	47571	-	10.0000	7471B	PPB	11 Mar 2020 09:03:46	S	Std
ICV53 - 1	ICV53	19423	4.0838	-	7471B	PPB	11 Mar 2020 09:06:29	U	SMPL
ICB53 - 1	ICB53	-252	-0.0522	-	7471B	PPB	11 Mar 2020 09:08:45	U	SMPL
CCV96 - 1	CCV96	24877	5.2303	-	7471B	PPB	11 Mar 2020 09:11:01	U	SMPL
CCB96 - 1	CCB96	-236	-0.0488	-	7471B	PPB	11 Mar 2020 09:13:16	U	SMPL
CRA - 1	CRA	920	0.1942	-	7471B	PPB	11 Mar 2020 09:15:33	U	SMPL
HighStd - 1	HighStd	48657	10.2292	-	7471B	PPB	11 Mar 2020 09:17:48	U	SMPL
ChkStd - 1	ChkStd	34916	7.3406	-	7471B	PPB	11 Mar 2020 09:20:03	U	SMPL
PB127434BL - 1	PBS	-330	-0.0686	-	7471B	PPB	11 Mar 2020 09:22:21	U	SMPL
PB127434BS - 1	LCSS	18644	3.9200	-	7471B	PPB	11 Mar 2020 09:24:39	U	SMPL
L1755-01 - 1	OK-03-031020	3945	0.8301	-	7471B	PPB	11 Mar 2020 09:26:55	U	SMPL
L1759-06 - 1	OR-03-031020	1621	0.3415	-	7471B	PPB	11 Mar 2020 09:29:13	U	SMPL
L1859-02 - 1	TP-7-SA	15460	3.2507	-	7471B	PPB	11 Mar 2020 09:31:30	U	SMPL
L1859-06 - 1	TP-7-DA	4622	0.9724	-	7471B	PPB	11 Mar 2020 09:33:46	U	SMPL
L1859-10 - 1	TP-8-SA	11693	2.4588	-	7471B	PPB	11 Mar 2020 09:36:02	U	SMPL
CCV97 - 1	CCV97	25374	5.3348	-	7471B	PPB	11 Mar 2020 09:38:17	U	SMPL
CCB97 - 1	CCB97	-23	-0.0041	-	7471B	PPB	11 Mar 2020 09:40:32	U	SMPL
L1864-01 - 1	SP-1	2429	0.5114	-	7471B	PPB	11 Mar 2020 09:42:48	U	SMPL
L1864-04 - 1	SP-2	1452	0.3060	-	7471B	PPB	11 Mar 2020 09:45:04	U	SMPL
L1864-07 - 1	SP-3	1350	0.2846	-	7471B	PPB	11 Mar 2020 09:47:19	U	SMPL
L1871-02 - 1	TP-17-SA	66904	14.0650	-	7471B	PPB	11 Mar 2020 09:49:35	U	SMPL
L1871-06 - 1	TP-17-DA	8895	1.8706	-	7471B	PPB	11 Mar 2020 09:51:51	U	SMPL
L1871-10 - 1	TP-18-SA	4250	0.8942	-	7471B	PPB	11 Mar 2020 09:54:09	U	SMPL
L1871-14 - 1	TP-18-DA	27927	5.8715	-	7471B	PPB	11 Mar 2020 09:56:25	U	SMPL
L1873-01 - 1	TP-1	758	0.1601	-	7471B	PPB	11 Mar 2020 09:58:42	U	SMPL
L1873-01DUP - 1	TP-1DUP	913	0.1927	-	7471B	PPB	11 Mar 2020 10:03:55	U	SMPL
L1873-01MS - 1	TP-1MS	20817	4.3768	-	7471B	PPB	11 Mar 2020 10:06:10	U	SMPL
CCV98 - 1	CCV98	25769	5.4178	-	7471B	PPB	11 Mar 2020 10:08:25	U	SMPL
CCB98 - 1	CCB98	-5	-0.0003	-	7471B	PPB	11 Mar 2020 10:10:41	U	SMPL
L1873-01MSD - 1	TP-1MSD	21305	4.4794	-	7471B	PPB	11 Mar 2020 10:12:57	U	SMPL
L1873-03 - 1	TP-2	1584	0.3338	-	7471B	PPB	11 Mar 2020 10:15:13	U	SMPL
L1873-05 - 1	TP-3	2741	0.5770	-	7471B	PPB	11 Mar 2020 10:17:29	U	SMPL
PB127435BL - 1	PBS	172	0.0369	-	7471B	PPB	11 Mar 2020 10:19:45	U	SMPL
PB127435BS - 1	LCSS	19258	4.0491	-	7471B	PPB	11 Mar 2020 10:22:00	U	SMPL
L1874-01 - 1	C-1	545	0.1153	-	7471B	PPB	11 Mar 2020 10:24:16	U	SMPL
L1874-01DUP - 1	C-1DUP	535	0.1132	-	7471B	PPB	11 Mar 2020 10:26:33	U	SMPL
L1874-01MS - 1	C-1MS	19013	3.9976	-	7471B	PPB	11 Mar 2020 10:28:50	U	SMPL
L1874-01MSD - 1	C-1MSD	19132	4.0226	-	7471B	PPB	11 Mar 2020 10:31:06	U	SMPL
L1874-03 - 1	C-2	514	0.1088	-	7471B	PPB	11 Mar 2020 10:33:23	U	SMPL
CCV99 - 1	CCV99	25639	5.3905	-	7471B	PPB	11 Mar 2020 10:35:39	U	SMPL
CCB99 - 1	CCB99	21	0.0052	-	7471B	PPB	11 Mar 2020 10:37:54	U	SMPL
L1874-05 - 1	C-3	723	0.1528	-	7471B	PPB	11 Mar 2020 10:40:11	U	SMPL
L1874-07 - 1	C-4	584	0.1235	-	7471B	PPB	11 Mar 2020 10:42:26	U	SMPL
L1874-09 - 1	C-5	683	0.1444	-	7471B	PPB	11 Mar 2020 10:44:41	U	SMPL
L1874-11 - 1	C-6	4025	0.8469	-	7471B	PPB	11 Mar 2020 10:46:57	U	SMPL
L1874-13 - 1	C-7	36550	7.6841	-	7471B	PPB	11 Mar 2020 10:49:12	U	SMPL
L1874-15 - 1	C-8	1480	0.3119	-	7471B	PPB	11 Mar 2020 10:54:03	U	SMPL
L1874-17 - 1	C-9	651	0.1376	-	7471B	PPB	11 Mar 2020 10:56:22	U	SMPL
CCV100 - 1	CCV100	23907	5.0264	-	7471B	PPB	11 Mar 2020 11:02:13	U	SMPL
CCB100 - 1	CCB100	2	0.0012	-	7471B	PPB	11 Mar 2020 11:04:30	U	SMPL
L1873-01LX5 - 1		268	0.0571	-	7471B	PPB	11 Mar 2020 11:06:46	U	SMPL
L1873-01A - 1		20993	4.4138	-	7471B	PPB	11 Mar 2020 11:09:01	U	SMPL
L1874-01LX5 - 1		134	0.0289	-	7471B	PPB	11 Mar 2020 11:11:17	U	SMPL
L1874-01A - 1		19260	4.0495	-	7471B	PPB	11 Mar 2020 11:13:33	U	SMPL
L1871-02DLX2 - 1		33297	7.0003	-	7471B	PPB	11 Mar 2020 11:15:48	U	SMPL
CCV01 - 1	CCV01	25239	5.3064	-	7471B	PPB	11 Mar 2020 11:18:04	U	SMPL
CCB01 - 1	CCB01	-16	-0.0026	-	7471B	PPB	11 Mar 2020 11:20:21	U	SMPL