



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108210

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/10/2020	11:10
2	CAL2	1	Water	NA	NA	20.2	7.00	03/10/2020	11:11
3	CAL3	1	Water	NA	NA	20.2	10.04	03/10/2020	11:12
4	ICV	1	Water	NA	NA	20.1	6.99	03/10/2020	11:15
5	CCV1	1	Water	NA	NA	20.2	2.01	03/10/2020	11:19
6	L1848-03	1	Water	NA	NA	19.3	5.66	03/10/2020	11:20
7	L1848-03DUP	1	Water	NA	NA	19.4	5.67	03/10/2020	11:21
8	CCV2	1	Water	NA	NA	20.2	12.02	03/10/2020	11:23

VB 108210

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1848 WorkList ID : 136885 Department : Wet-Chemistry Date : 03-10-2020 10:45:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/05/2020	Water	L1848-03	pH	Cool 4 deg C	STER15	D11	GILROY-OMEGA-CELL	03/05/2020	SM 4500-H B

Date/Time 03-10-2020 11:00
 Received by: JN
 Relinquished by: JN

Date/Time 03-10-2020 12:15
 Received by: JP
 Relinquished by: JN