



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108220

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/11/2020	16:15
2	CAL2	1	Water	NA	NA	20.1	6.99	03/11/2020	16:15
3	CAL3	1	Water	NA	NA	20.1	10.03	03/11/2020	16:16
4	ICV	1	Water	NA	NA	20.1	7.00	03/11/2020	16:19
5	CCV1	1	Water	NA	NA	20.2	2.01	03/11/2020	16:20
6	L1887-05	1	Water	NA	NA	20.1	6.62	03/11/2020	16:21
7	L1887-05DUP	1	Water	NA	NA	20.2	6.63	03/11/2020	16:22
8	CCV2	1	Water	NA	NA	20.1	12.02	03/11/2020	16:25

2B108220

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L1887 WorkList ID : 136932 Department : Wet-Chemistry Date : 03-11-2020 16:03:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/11/2020	Water	L1887-05	Corrosivity	Cool 4 deg C	PSEG01	I53	CL-GW	03/11/2020	9040C

Date/Time 03-11-2020 16:10
Received by: JR
Relinquished by: JR

Date/Time 03-11-2020 17:15
Received by: JR
Relinquished by: JR