



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108221

Slope : 98.8

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2628         |
| buffer solution pH 7 yellow    | W2626         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2632         |
| buffer solution pH 7 yellow    | W2569         |
| Buffer Solution, PH2 (500ml)   | W2592         |
| Buffer Solution, PH12 (500ml)  | W2616         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.1             | 4.01        | 03/11/2020 | 16:15     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.1             | 6.99        | 03/11/2020 | 16:15     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.1             | 10.03       | 03/11/2020 | 16:16     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.1             | 7.00        | 03/11/2020 | 16:19     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.2             | 2.01        | 03/11/2020 | 16:20     |
| 6   | L1887-04    | 1  | Water  | NA          | NA          | 20.1             | 6.62        | 03/11/2020 | 16:21     |
| 7   | L1887-04DUP | 1  | Water  | NA          | NA          | 20.2             | 6.63        | 03/11/2020 | 16:22     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.1             | 12.02       | 03/11/2020 | 16:25     |

# WORKLIST(Hardcopy Internal Chain)

283108221

WorkList Name : PH L1887

WorkList ID : 136933

Department : Wet-Chemistry

Date : 03-11-2020 16:04:01

| Due Date   | Matrix | Sample                 | Test | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|------------------------|------|--------------|----------|------------------|-----------------|--------------|--------|
| 03/11/2020 | Water  | L1887-04- <sup>4</sup> | pH   | Cool 4 deg C | PSEG01   | I53              | CL-GW           | 03/11/2020   | 9040C  |

Date/Time 03.11.2020 16:10

Received by: JR

Relinquished by: JR

Date/Time 03.11.2020 17:15

Received by: JR

Relinquished by: JR