



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 3/12/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 03/11/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 102
Time OUT: 08:00
Out Date: 03/12/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB108223

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L1752-01	HD-01-031120	1	1.15	9.68	8.89	90.7	
L1752-02	HD-01-031120	2	1.13	9.88	9.11	91.2	
L1752-03	HD-01-031120-1	3	1.13	9.69	9.4	96.6	
L1876-01	400310006	4	1.16	9.64	1.56	4.7	SLUDGE SAMPLE
L1877-01	400305803	5	1.15	9.95	2.76	18.3	SLUDGE SAMPLE
L1878-01	400305023	6	1.19	9.97	1.47	3.2	SLUDGE SAMPLE
L1879-01	400306038	7	1.15	9.97	1.55	4.5	SLUDGE SAMPLE
L1880-01	400306700	8	1.14	9.92	3.4	25.7	SLUDGE SAMPLE
L1881-01	400305749	9	1.17	9.56	2.72	18.5	SLUDGE SAMPLE
L1882-01	PCB-110819-01	10	1.00	2.00	2.00	100.0	CAULKING SAMPLE
L1882-02	PCB-110819-02	11	1.00	2.00	2.00	100.0	CAULKING SAMPLE
L1887-01	CL-TP	12	1.12	9.6	8.65	88.8	
L1887-02	CL-TP-B	13	1.15	9.84	9.00	90.3	
L1889-01	CP-1	14	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100% SOLIDS
L1889-02	CP-2	15	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100% SOLIDS
L1889-03	CP-3	16	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100% SOLIDS
L1889-04	CP-4	17	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100% SOLIDS
L1889-05	CP-5	18	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100% SOLIDS
L1889-06	CP-6	19	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100% SOLIDS
L1893-01	TP-1 (0-1)	20	1.13	9.57	8.34	85.4	
L1893-02	TP-1-SA	21	1.17	9.62	8.08	81.8	
L1893-03	TP-1-SB	22	1.15	9.94	8.46	83.2	
L1893-05	TP-2 (2-3)	23	1.12	9.53	8.5	87.8	
L1893-06	TP-2-SA	24	1.11	9.8	8.63	86.5	
L1893-07	TP-2-SB	25	1.16	9.66	8.51	86.5	
L1893-09	TP-2 (8-9)	26	1.12	9.82	8.58	85.7	
L1893-10	TP-2-DA	27	1.18	9.97	8.7	85.6	
L1893-11	TP-2-DB	28	1.1	9.96	8.71	85.9	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 3/12/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 03/11/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 102
Time OUT: 08:00
Out Date: 03/12/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB108223

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L1893-14	TP-4-SA	29	1.17	9.87	8.51	84.4	
L1893-15	TP-4-SB	30	1.18	9.63	8.5	86.6	
L1893-17	TP-14 (1.5-2)	31	1.15	9.88	8.58	85.1	
L1893-18	TP-14-SA	32	1.13	9.62	8.24	83.7	
L1893-19	TP-14-SB	33	1.13	9.55	7.81	79.3	
L1893-21	TP-14 (8-9)	34	1.12	9.9	8.22	80.9	
L1893-22	TP-14-DA	35	1.15	9.76	8.1	80.7	
L1893-23	TP-14-DB	36	1.17	9.96	8.66	85.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$