



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108224

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/11/2020	17:25
2	CAL2	1	Water	NA	NA	20.1	7.01	03/11/2020	17:26
3	CAL3	1	Water	NA	NA	20.1	10.03	03/11/2020	17:27
4	ICV	1	Water	NA	NA	20.1	7.00	03/11/2020	17:30
5	CCV1	1	Water	NA	NA	20.2	2.01	03/11/2020	17:32
6	L1752-04	1	Solid	20.03	20	20.5	9.23	03/11/2020	17:35
7	L1752-04DUP	1	Solid	20.04	20	20.6	9.24	03/11/2020	17:36
8	L1887-03	1	Solid	20.02	20	20.2	7.96	03/11/2020	17:38
9	L1893-04	1	Solid	20.04	20	20.5	6.99	03/11/2020	17:40
10	L1893-08	1	Solid	20.02	20	20.2	8.54	03/11/2020	17:42
11	L1893-12	1	Solid	20.03	20	20.4	7.51	03/11/2020	17:45
12	L1893-16	1	Solid	20.05	20	20.2	7.37	03/11/2020	17:46
13	L1893-20	1	Solid	20.02	20	20.7	7.20	03/11/2020	17:48
14	L1893-24	1	Solid	20.03	20	20.6	7.66	03/11/2020	17:50
15	CCV2	1	Water	NA	NA	20.2	12.02	03/11/2020	18:00

203108224

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1892 WorkList ID : 136931 Department : Wet-Chemistry Date : 03-11-2020 16:01:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/11/2020	Solid	L1752-04	Corrosivity	Cool 4 deg C	PSEG05	I53	HD-01-031120	03/11/2020	9045D
03/11/2020	Solid	L1887-03	Corrosivity	Cool 4 deg C	PSEG01	I53	CL-TP	03/11/2020	9045D
03/10/2020	Solid	L1893-04	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-1-SA	03/10/2020	9045D
03/10/2020	Solid	L1893-08	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-2-SA	03/10/2020	9045D
03/10/2020	Solid	L1893-12	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-2-DA	03/10/2020	9045D
03/10/2020	Solid	L1893-16	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-4-SA	03/10/2020	9045D
03/10/2020	Solid	L1893-20	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-14-SA	03/10/2020	9045D
03/10/2020	Solid	L1893-24	Corrosivity	Cool 4 deg C	PSEG01	I53	TP-14-DA	03/10/2020	9045D

Date/Time 03-11-2020 16:00
 Received by: Jn
 Relinquished by: Jn

Date/Time 03-11-2020 18:00
 Received by: Dk
 Relinquished by: Jn