


LB108233

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.064	0.064	363	0.000	363				
0.2	0.200	0.237	0.037	984	0.0 %	984				
2.5	2.500	2.553	0.053	9309	0.0 %	9309				
5.0	5.000	4.953	-0.047	17937	0.0 %	17937				
7.5	7.500	7.125	-0.375	25747	0.0 %	25747				
10.0	10.000	10.268	0.268	37045	0.0 %	37045				

LB108233 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Mar 2020 09:33:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	363	-	0.0000	7470A	PPB	12 Mar 2020 09:43:20	S	Std
0.2 - 1	50.2	984	-	0.2000	7470A	PPB	12 Mar 2020 09:45:28	S	Std
2.5 - 1	52.5	9309	-	2.5000	7470A	PPB	12 Mar 2020 09:47:35	S	Std
5.0 - 1	55	17937	-	5.0000	7470A	PPB	12 Mar 2020 09:49:42	S	Std
7.5 - 1	57.5	25747	-	7.5000	7470A	PPB	12 Mar 2020 09:51:50	S	Std
10.0 - 1	57.0	37045	-	10.0000	7470A	PPB	12 Mar 2020 09:53:57	S	Std
ICV55 - 1	ICV55	14041	3.8690	-	7470A	PPB	12 Mar 2020 09:56:38	U	SMPL
ICB55 - 1	ICB55	137	0.0011	-	7470A	PPB	12 Mar 2020 09:58:43	U	SMPL
CCV08 - 1	CCV08	18490	5.1066	-	7470A	PPB	12 Mar 2020 10:00:50	U	SMPL
CCB08 - 1	CCB08	97	-0.0100	-	7470A	PPB	12 Mar 2020 10:02:56	U	SMPL
CRA - 1	CRA	907	0.2153	-	7470A	PPB	12 Mar 2020 10:05:02	U	SMPL
HighStd - 1	HighStd	37635	10.4325	-	7470A	PPB	12 Mar 2020 10:07:08	U	SMPL
ChkStd - 1	ChkStd	26275	7.2723	-	7470A	PPB	12 Mar 2020 10:09:13	U	SMPL
PB127456BL - 1	PBW	-37	-0.0473	-	7470A	PPB	12 Mar 2020 10:11:21	U	SMPL
PB127456BS - 1	LCSW	14063	3.8751	-	7470A	PPB	12 Mar 2020 10:13:29	U	SMPL
L1755-04 - 1	OK-03-031020	104	-0.0080	-	7470A	PPB	12 Mar 2020 10:15:37	U	SMPL
L1755-04DUP - 1	OK-03-031020DUP	171	0.0106	-	7470A	PPB	12 Mar 2020 10:22:14	U	SMPL
L1755-04MS - 1	OK-03-031020MS	13996	3.8565	-	7470A	PPB	12 Mar 2020 10:24:21	U	SMPL
L1755-04MSD - 1	OK-03-031020MSD	13757	3.7900	-	7470A	PPB	12 Mar 2020 10:26:29	U	SMPL
L1759-10 - 1	OR-03-031020	135	0.0006	-	7470A	PPB	12 Mar 2020 10:28:36	U	SMPL
CCV09 - 1	CCV09	18288	5.0505	-	7470A	PPB	12 Mar 2020 10:30:43	U	SMPL
CCB09 - 1	CCB09	108	-0.0069	-	7470A	PPB	12 Mar 2020 10:32:48	U	SMPL
L1853-04 - 1	CB-EC-52-G	285	0.0423	-	7470A	PPB	12 Mar 2020 10:34:55	U	SMPL
L1853-07 - 1	CB-EC-52-S	357	0.0623	-	7470A	PPB	12 Mar 2020 10:37:01	U	SMPL
L1853-10 - 1	CB-EC-21-G	341	0.0579	-	7470A	PPB	12 Mar 2020 10:39:06	U	SMPL
L1853-13 - 1	CB-EC-21-S	333	0.0557	-	7470A	PPB	12 Mar 2020 10:41:13	U	SMPL
L1859-04 - 1	TP-7-SA	353	0.0612	-	7470A	PPB	12 Mar 2020 10:43:20	U	SMPL
L1859-08 - 1	TP-7-DA	343	0.0584	-	7470A	PPB	12 Mar 2020 10:45:27	U	SMPL
L1859-12 - 1	TP-8-SA	370	0.0660	-	7470A	PPB	12 Mar 2020 10:47:34	U	SMPL
L1864-02 - 1	SP-1	361	0.0635	-	7470A	PPB	12 Mar 2020 10:49:41	U	SMPL
L1864-05 - 1	SP-2	322	0.0526	-	7470A	PPB	12 Mar 2020 10:51:49	U	SMPL
L1873-02 - 1	TP-1	387	0.0707	-	7470A	PPB	12 Mar 2020 10:53:56	U	SMPL
CCV10 - 1	CCV10	18209	5.0285	-	7470A	PPB	12 Mar 2020 10:56:01	U	SMPL
CCB10 - 1	CCB10	133	0.0000	-	7470A	PPB	12 Mar 2020 10:58:07	U	SMPL
L1873-04 - 1	TP-2	320	0.0520	-	7470A	PPB	12 Mar 2020 11:00:15	U	SMPL
L1873-06 - 1	TP-3	360	0.0632	-	7470A	PPB	12 Mar 2020 11:02:21	U	SMPL
PB127457BL - 1	PBW	337	0.0568	-	7470A	PPB	12 Mar 2020 11:04:27	U	SMPL
PB127457BS - 1	LCSW	13640	3.7575	-	7470A	PPB	12 Mar 2020 11:06:33	U	SMPL
L1871-04 - 1	TP-17-SA	199	0.0184	-	7470A	PPB	12 Mar 2020 11:08:40	U	SMPL
L1871-04DUP - 1	TP-17-SADUP	241	0.0301	-	7470A	PPB	12 Mar 2020 11:13:20	U	SMPL
L1871-04MS - 1	TP-17-SAMS	14938	4.1185	-	7470A	PPB	12 Mar 2020 11:15:27	U	SMPL
L1871-04MSD - 1	TP-17-SAMSD	13824	3.8086	-	7470A	PPB	12 Mar 2020 11:17:34	U	SMPL
L1871-08 - 1	TP-17-DA	289	0.0434	-	7470A	PPB	12 Mar 2020 11:19:42	U	SMPL
L1871-12 - 1	TP-18-SA	356	0.0621	-	7470A	PPB	12 Mar 2020 11:21:49	U	SMPL
CCV11 - 1	CCV11	17898	4.9420	-	7470A	PPB	12 Mar 2020 11:23:55	U	SMPL
CCB11 - 1	CCB11	160	0.0075	-	7470A	PPB	12 Mar 2020 11:26:01	U	SMPL
L1871-16 - 1	TP-18-DA	338	0.0571	-	7470A	PPB	12 Mar 2020 11:28:08	U	SMPL
L1874-02 - 1	C-1	333	0.0557	-	7470A	PPB	12 Mar 2020 11:30:14	U	SMPL
L1874-04 - 1	C-2	359	0.0629	-	7470A	PPB	12 Mar 2020 11:32:20	U	SMPL
L1874-06 - 1	C-3	379	0.0685	-	7470A	PPB	12 Mar 2020 11:34:26	U	SMPL
L1874-08 - 1	C-4	356	0.0621	-	7470A	PPB	12 Mar 2020 11:36:33	U	SMPL
L1874-10 - 1	C-5	355	0.0618	-	7470A	PPB	12 Mar 2020 11:38:40	U	SMPL
L1874-12 - 1	C-6	369	0.0657	-	7470A	PPB	12 Mar 2020 11:40:47	U	SMPL
L1874-14 - 1	C-7	410	0.0771	-	7470A	PPB	12 Mar 2020 11:42:55	U	SMPL
L1874-16 - 1	C-8	383	0.0696	-	7470A	PPB	12 Mar 2020 11:45:03	U	SMPL
L1874-18 - 1	C-9	376	0.0676	-	7470A	PPB	12 Mar 2020 11:47:09	U	SMPL
CCV12 - 1	CCV12	18037	4.9806	-	7470A	PPB	12 Mar 2020 11:49:15	U	SMPL
CCB12 - 1	CCB12	136	0.0009	-	7470A	PPB	12 Mar 2020 11:51:22	U	SMPL
PB127456TB - 1	PB127430TB	435	0.0840	-	7470A	PPB	12 Mar 2020 11:53:29	U	SMPL
PB127457TB - 1	PB127455TB	407	0.0762	-	7470A	PPB	12 Mar 2020 11:55:36	U	SMPL
L1871-04LX5 - 1		322	0.0526	-	7470A	PPB	12 Mar 2020 11:57:42	U	SMPL
L1871-04A - 1		14170	3.9049	-	7470A	PPB	12 Mar 2020 11:59:48	U	SMPL
L1755-04LX5 - 1		191	0.0162	-	7470A	PPB	12 Mar 2020 12:01:54	U	SMPL
L1755-04A - 1		14176	3.9066	-	7470A	PPB	12 Mar 2020 12:04:01	U	SMPL
CCV13 - 1	CCV13	18760	5.1818	-	7470A	PPB	12 Mar 2020 12:06:09	U	SMPL
CCB13 - 1	CCB13	133	0.0000	-	7470A	PPB	12 Mar 2020 12:08:16	U	SMPL