


LB108239

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	81	0.000	81				
0.2	0.200	0.171	-0.029	982	0.0 %	982				
2.5	2.500	2.410	-0.090	12328	0.0 %	12328				
5.0	5.000	5.115	0.115	26034	0.0 %	26034				
7.5	7.500	7.733	0.233	39302	0.0 %	39302				
10.0	10.000	9.779	-0.221	49668	0.0 %	49668				

LB108239 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Mar 2020 12:42:04

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	81	-	0.00007471B	PPB	12 Mar 2020 12:43:08	S	Std
0.2 - 1	50.2	982	-	0.20007471B	PPB	12 Mar 2020 12:45:25	S	Std
2.5 - 1	52.5	12328	-	2.50007471B	PPB	12 Mar 2020 12:47:42	S	Std
5.0 - 1	55	26034	-	5.00007471B	PPB	12 Mar 2020 12:49:59	S	Std
7.5 - 1	57.5	39302	-	7.50007471B	PPB	12 Mar 2020 12:52:16	S	Std
10.0 - 1	510	49668	-	10.00007471B	PPB	12 Mar 2020 12:54:33	S	Std
ICV56 - 1	ICV56	19825	3.8894	-7471B	PPB	12 Mar 2020 12:57:45	U	SMPL
ICB56 - 1	ICB56	-154	-0.0533	-7471B	PPB	12 Mar 2020 12:59:59	U	SMPL
CCV14 - 1	CCV14	26034	5.1146	-7471B	PPB	12 Mar 2020 13:02:16	U	SMPL
CCB14 - 1	CCB14	-195	-0.0614	-7471B	PPB	12 Mar 2020 13:04:31	U	SMPL
CRA - 1	CRA	940	0.1626	-7471B	PPB	12 Mar 2020 13:06:47	U	SMPL
HighStd - 1	HighStd	49843	9.8131	-7471B	PPB	12 Mar 2020 13:09:02	U	SMPL
ChkStd - 1	ChkStd	35568	6.9961	-7471B	PPB	12 Mar 2020 13:11:17	U	SMPL
PB127468BL - 1	PBS	-441	-0.1099	-7471B	PPB	12 Mar 2020 13:13:35	U	SMPL
PB127468BS - 1	LCSS	17898	3.5091	-7471B	PPB	12 Mar 2020 13:15:53	U	SMPL
L1752-01 - 1	HD-01-031120	5190	1.0013	-7471B	PPB	12 Mar 2020 13:18:09	U	SMPL
L1887-01 - 1	CL-TP	4129	0.7919	-7471B	PPB	12 Mar 2020 13:20:27	U	SMPL
L1893-02 - 1	TP-1-SA	43289	8.5197	-7471B	PPB	12 Mar 2020 13:22:44	U	SMPL
L1893-06 - 1	TP-2-SA	28406	5.5827	-7471B	PPB	12 Mar 2020 13:25:01	U	SMPL
L1893-10 - 1	TP-2-DA	8075	1.5706	-7471B	PPB	12 Mar 2020 13:27:18	U	SMPL
CCV15 - 1	CCV15	26449	5.1965	-7471B	PPB	12 Mar 2020 13:29:35	U	SMPL
CCB15 - 1	CCB15	-78	-0.0383	-7471B	PPB	12 Mar 2020 13:31:50	U	SMPL
L1893-14 - 1	TP-4-SA	4094	0.7850	-7471B	PPB	12 Mar 2020 13:39:08	U	SMPL
L1893-14DUP - 1	TP-4-SADUP	3904	0.7475	-7471B	PPB	12 Mar 2020 13:41:23	U	SMPL
L1893-14MS - 1	TP-4-SAMS	25001	4.9108	-7471B	PPB	12 Mar 2020 13:43:38	U	SMPL
L1893-14MSD - 1	TP-4-SAMSD	24635	4.8386	-7471B	PPB	12 Mar 2020 13:45:53	U	SMPL
L1893-18 - 1	TP-14-SA	3761	0.7193	-7471B	PPB	12 Mar 2020 13:48:10	U	SMPL
L1893-22 - 1	TP-14-DA	7380	1.4335	-7471B	PPB	12 Mar 2020 13:50:27	U	SMPL
L1893-14LX5 - 1		721	0.1194	-7471B	PPB	12 Mar 2020 13:52:44	U	SMPL
L1893-14A - 1		24885	4.8879	-7471B	PPB	12 Mar 2020 13:55:01	U	SMPL
CCV16 - 1	CCV16	27080	5.3211	-7471B	PPB	12 Mar 2020 13:57:17	U	SMPL
CCB16 - 1	CCB16	-60	-0.0347	-7471B	PPB	12 Mar 2020 13:59:33	U	SMPL