



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB108240

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	03/12/2020	09:50
2	L1887-05		1	9	100.00	>212.0	>212.0	03/12/2020	10:25
3	L1887-05DUP		1	9	100.00	>212.0	>212.0	03/12/2020	10:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp31220

WorkList ID : 136957

Department : Wet-Chemistry

Date : 03-12-2020 09:14:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/16/2020	Water	L1887-05	C	Cool 4 deg C	PSEG01	I53	CL-GW	03/11/2020	1010A

Date/Time 03/12/2020 09:20
 Received by: KS
 Relinquished by: JS

Date/Time 03/12/20 11:00
 Received by: JS
 Relinquished by: KS