

LB108241


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4279E



A= 0.0000e+000

B= 2.3506e-004

C= -1.2664e-002

Rho= 0.9999086

Accept=Accepted

slope
intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.003	0.003	68	0.000	68					-
0.05	0.050					149					-
0.20	0.200	0.191	-0.009	868	0.0 %	868					5
2.50	2.500	2.509	0.009	10728	0.0 %	10728					0
5.00	5.000	5.050	0.050	21537	0.0 %	21537					1
7.50	7.500	7.400	-0.100	31533	0.0 %	31533					-1
10.0	10.000	10.047	0.047	42795	0.0 %	42795					0

LB108241
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	68	-		0 ISM02.4	PPB	3/12/2020 15:13	Std
	0.2 S0.2	868	-		0.2 ISM02.4	PPB	3/12/2020 15:15	Std
	2.5 S2.5	10728	-		2.5 ISM02.4	PPB	3/12/2020 15:18	Std
	5 S5.0	21537	-		5 ISM02.4	PPB	3/12/2020 15:20	Std
	7.5 S7.5	31533	-		7.5 ISM02.4	PPB	3/12/2020 15:22	Std
	10 S10.0	42795	-		10 ISM02.4	PPB	3/12/2020 15:24	Std
ICV	ICV	16695	3.9117 -		ISM02.4	PPB	3/12/2020 15:27	SMPL
ICB	ICB	-95	-0.035 -		ISM02.4	PPB	3/12/2020 15:29	SMPL
CCV017	CCV017	21730	5.0952 -		ISM02.4	PPB	3/12/2020 15:32	SMPL
CCB017	CCB017	-104	-0.0371 -		ISM02.4	PPB	3/12/2020 15:34	SMPL
PB127494BL	PBW494	7	-0.011 -		ISM02.4	PPB	3/12/2020 15:36	SMPL
L1894-01	MDBERO	-4	-0.0136 -		ISM02.4	PPB	3/12/2020 15:38	SMPL
L1894-02	MDBEROD	20	-0.008 -		ISM02.4	PPB	3/12/2020 15:41	SMPL
L1894-03	MDBEROS	4218	0.9788 -		ISM02.4	PPB	3/12/2020 15:43	SMPL
CCV018	CCV018	21086	4.9439 -		ISM02.4	PPB	3/12/2020 15:45	SMPL
CCB018	CCB018	-142	-0.046 -		ISM02.4	PPB	3/12/2020 15:48	SMPL