

LB 108243


ISM02.4

INSTRUMENT ID : CV 1

Linear

 μ Abs.:

4226E



A= 0.0000e+000

B= 2.3841e-004

C= 3.0002e-003

Rho= 0.9998942

Accept=Accepted

slope
intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.018	0.018	63	0.000	63					1
0.05	0.050					149					1
0.20	0.200	0.206	0.006	851	0.0 %	851					3
2.50	2.500	2.525	0.025	10579	0.0 %	10579					1
5.00	5.000	4.965	-0.035	20813	0.0 %	20813					1
7.50	7.500	7.406	-0.094	31054	0.0 %	31054					1
10.0	10.000	10.080	0.080	42266	0.0 %	42266					1

LB108243

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	63	-		0 ISM02.4	PPB	3/12/2020 16:03	Std
	0.2 S0.2	851	-		0.2 ISM02.4	PPB	3/12/2020 16:06	Std
	2.5 S2.5	10579	-		2.5 ISM02.4	PPB	3/12/2020 16:08	Std
	5 S5.0	20813	-		5 ISM02.4	PPB	3/12/2020 16:10	Std
	7.5 S7.5	31054	-		7.5 ISM02.4	PPB	3/12/2020 16:13	Std
	10 S10.0	42266	-		10 ISM02.4	PPB	3/12/2020 16:15	Std
ICV	ICV	15371	3.6676 -		ISM02.4	PPB	3/12/2020 16:18	SMPL
ICB	ICB	-131	-0.0282 -		ISM02.4	PPB	3/12/2020 16:20	SMPL
CCV019	CCV019	21012	5.0124 -		ISM02.4	PPB	3/12/2020 16:22	SMPL
CCB019	CCB019	-132	-0.0285 -		ISM02.4	PPB	3/12/2020 16:24	SMPL
PB127495BL	PBW495	7	0.0047 -		ISM02.4	PPB	3/12/2020 16:27	SMPL
L1895-01	MDBER1	-54	-0.0099 -		ISM02.4	PPB	3/12/2020 16:31	SMPL
L1895-02	MDBER2	28	0.0097 -		ISM02.4	PPB	3/12/2020 16:33	SMPL
L1895-03	MDBER2D	14	0.0063 -		ISM02.4	PPB	3/12/2020 16:36	SMPL
L1895-04	MDBER2S	4568	1.092 -		ISM02.4	PPB	3/12/2020 16:38	SMPL
PB127495TB	LEB495	405	0.0996 -		ISM02.4	PPB	3/12/2020 16:40	SMPL
CCV020	CCV020	20127	4.8014 -		ISM02.4	PPB	3/12/2020 16:42	SMPL
CCB020	CCB020	-112	-0.0237 -		ISM02.4	PPB	3/12/2020 16:45	SMPL