



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108250

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/12/2020	17:25
2	CAL2	1	Water	NA	NA	20.1	7.01	03/12/2020	17:26
3	CAL3	1	Water	NA	NA	20.2	10.03	03/12/2020	17:28
4	ICV	1	Water	NA	NA	20.1	7.00	03/12/2020	17:30
5	CCV1	1	Water	NA	NA	20.2	2.02	03/12/2020	17:50
6	L1912-01	1	Solid	20.03	20	21.0	7.50	03/12/2020	18:00
7	L1912-01DUP	1	Solid	20.04	20	21.1	7.51	03/12/2020	18:01
8	L1912-05	1	Solid	20.02	20	20.9	8.53	03/12/2020	18:02
9	L1912-09	1	Solid	20.03	20	20.8	8.62	03/12/2020	18:05
10	L1912-13	1	Solid	20.04	20	20.7	7.31	03/12/2020	18:06
11	L1912-17	1	Solid	20.02	20	20.5	7.28	03/12/2020	18:10
12	L1912-21	1	Solid	20.04	20	20.7	7.06	03/12/2020	18:11
13	L1912-25	1	Solid	20.03	20	20.8	6.24	03/12/2020	18:12
14	L1912-29	1	Solid	20.04	20	20.8	7.36	03/12/2020	18:15
15	L1914-01	1	Solid	20.02	20	20.8	6.90	03/12/2020	18:20
16	CCV2	1	Water	NA	NA	20.1	12.02	03/12/2020	18:21

VB 108250

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1912 WorkList ID : 136978 Department : Wet-Chemistry Date : 03-12-2020 13:40:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/12/2020	Solid	L1912-01	pH	Cool 4 deg C	PSEG01	I61	Z1-1	03/12/2020	9045D
03/12/2020	Solid	L1912-05	pH	Cool 4 deg C	PSEG01	I61	Z3-S1	03/12/2020	9045D
03/12/2020	Solid	L1912-09	pH	Cool 4 deg C	PSEG01	I61	Z3-S2	03/12/2020	9045D
03/12/2020	Solid	L1912-13	pH	Cool 4 deg C	PSEG01	I61	Z1-2	03/12/2020	9045D
03/12/2020	Solid	L1912-17	pH	Cool 4 deg C	PSEG01	I61	Z1-3	03/12/2020	9045D
03/12/2020	Solid	L1912-21	pH	Cool 4 deg C	PSEG01	I61	Z1-4	03/12/2020	9045D
03/12/2020	Solid	L1912-25	pH	Cool 4 deg C	PSEG01	I61	Z1-5	03/12/2020	9045D
03/12/2020	Solid	L1912-29	pH	Cool 4 deg C	PSEG01	I61	Z1-6	03/12/2020	9045D
03/12/2020	Solid	L1914-01	pH	Cool 4 deg C	PSEG01	I61	32292-32293	03/12/2020	9045D

Date/Time 03.12.2020 16.00
 Received by: Jo
 Relinquished by: K.C

Date/Time 03.12.2020 18:10
 Received by: KC
 Relinquished by: Jh