


LB108254

7470A

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	-19	0.000	-19				
0.2	0.200	0.193	-0.007	697	0.0 %	697				
2.5	2.500	2.497	-0.003	8838	0.0 %	8838				
5.0	5.000	4.994	-0.006	17660	0.0 %	17660				
7.5	7.500	7.597	0.097	26855	0.0 %	26855				
10.0	10.000	9.930	-0.070	35098	0.0 %	35098				

LB108254 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Mar 2020 08:56:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-19	-	0.0000	7470A	PPB	13 Mar 2020 08:59:37	S	Std
0.2 - 1	50.2	697	-	0.2000	7470A	PPB	13 Mar 2020 09:01:44	S	Std
2.5 - 1	52.5	8838	-	2.5000	7470A	PPB	13 Mar 2020 09:03:51	S	Std
5.0 - 1	55	17660	-	5.0000	7470A	PPB	13 Mar 2020 09:05:59	S	Std
7.5 - 1	57.5	26855	-	7.5000	7470A	PPB	13 Mar 2020 09:08:06	S	Std
10.0 - 1	510	35098	-	10.0000	7470A	PPB	13 Mar 2020 09:10:13	S	Std
ICV57 - 1	ICV57	13145	3.7160	-7470A	PPB	PPB	13 Mar 2020 09:12:48	U	SMPL
ICB57 - 1	ICB57	-109	-0.0356	-7470A	PPB	PPB	13 Mar 2020 09:17:18	U	SMPL
CCV21 - 1	CCV21	17549	4.9626	-7470A	PPB	PPB	13 Mar 2020 09:21:44	U	SMPL
CCB21 - 1	CCB21	-297	-0.0888	-7470A	PPB	PPB	13 Mar 2020 09:23:49	U	SMPL
CRA - 1	CRA	605	0.1665	-7470A	PPB	PPB	13 Mar 2020 09:25:56	U	SMPL
HighStd - 1	HighStd	35885	10.1527	-7470A	PPB	PPB	13 Mar 2020 09:28:02	U	SMPL
ChkStd - 1	ChkStd	25110	7.1028	-7470A	PPB	PPB	13 Mar 2020 09:30:08	U	SMPL
PB127486BL - 1	PBW	-117	-0.0379	-7470A	PPB	PPB	13 Mar 2020 09:34:45	U	SMPL
PB127486BS - 1	LCSW	12359	3.4935	-7470A	PPB	PPB	13 Mar 2020 09:36:52	U	SMPL
L1752-04 - 1	HD-01-031120	-239	-0.0724	-7470A	PPB	PPB	13 Mar 2020 09:38:59	U	SMPL
L1752-04DUP - 1	HD-01-031120DUP	-89	-0.0299	-7470A	PPB	PPB	13 Mar 2020 09:41:07	U	SMPL
L1752-04MS - 1	HD-01-031120MS	12061	3.4092	-7470A	PPB	PPB	13 Mar 2020 09:43:14	U	SMPL
L1752-04MSD - 1	HD-01-031120MSD	13065	3.6934	-7470A	PPB	PPB	13 Mar 2020 09:45:21	U	SMPL
L1797-03 - 1	6-FEET	-259	-0.0781	-7470A	PPB	PPB	13 Mar 2020 09:47:27	U	SMPL
CCV22 - 1	CCV22	18420	5.2091	-7470A	PPB	PPB	13 Mar 2020 09:49:33	U	SMPL
CCB22 - 1	CCB22	-256	-0.0772	-7470A	PPB	PPB	13 Mar 2020 09:51:39	U	SMPL
L1797-04 - 1	30-FEET	-86	-0.0291	-7470A	PPB	PPB	13 Mar 2020 09:53:46	U	SMPL
L1887-03 - 1	CL-TP	-58	-0.0212	-7470A	PPB	PPB	13 Mar 2020 09:55:52	U	SMPL
L1893-04 - 1	TP-1-SA	8	-0.0025	-7470A	PPB	PPB	13 Mar 2020 09:57:58	U	SMPL
L1893-08 - 1	TP-2-SA	-10	-0.0076	-7470A	PPB	PPB	13 Mar 2020 10:00:05	U	SMPL
L1893-12 - 1	TP-2-DA	-2	-0.0053	-7470A	PPB	PPB	13 Mar 2020 10:02:11	U	SMPL
L1893-16 - 1	TP-4-SA	-5	-0.0062	-7470A	PPB	PPB	13 Mar 2020 10:04:18	U	SMPL
L1893-20 - 1	TP-14-SA	190	0.0490	-7470A	PPB	PPB	13 Mar 2020 10:06:25	U	SMPL
L1893-24 - 1	TP-14-DA	0	-0.0047	-7470A	PPB	PPB	13 Mar 2020 10:08:33	U	SMPL
PB127487BL - 1	PBW	-11	-0.0079	-7470A	PPB	PPB	13 Mar 2020 10:10:41	U	SMPL
PB127487BS - 1	LCSW	13212	3.7350	-7470A	PPB	PPB	13 Mar 2020 10:12:47	U	SMPL
CCV23 - 1	CCV23	17822	5.0399	-7470A	PPB	PPB	13 Mar 2020 10:14:53	U	SMPL
CCB23 - 1	CCB23	-180	-0.0557	-7470A	PPB	PPB	13 Mar 2020 10:17:00	U	SMPL
L1887-04 - 1	CL-GW	1810	0.5076	-7470A	PPB	PPB	13 Mar 2020 10:21:56	U	SMPL
L1887-04DUP - 1	CL-GWDUP	1531	0.4286	-7470A	PPB	PPB	13 Mar 2020 10:24:02	U	SMPL
L1887-04MS - 1	CL-GWMS	17449	4.9343	-7470A	PPB	PPB	13 Mar 2020 10:26:08	U	SMPL
L1887-04MSD - 1	CL-GWMSD	18022	5.0965	-7470A	PPB	PPB	13 Mar 2020 10:28:14	U	SMPL
L1913-19 - 1	FIELD-BLANK	-91	-0.0305	-7470A	PPB	PPB	13 Mar 2020 10:30:20	U	SMPL
PB127515BL - 1	PBW	10	-0.0019	-7470A	PPB	PPB	13 Mar 2020 10:32:26	U	SMPL
PB127515BS - 1	LCSW	13422	3.7944	-7470A	PPB	PPB	13 Mar 2020 10:34:33	U	SMPL
L1887-05 - 1	CL-GW	-53	-0.0197	-7470A	PPB	PPB	13 Mar 2020 10:36:41	U	SMPL
L1887-05DUP - 1	CL-GWDUP	14	-0.0008	-7470A	PPB	PPB	13 Mar 2020 10:38:49	U	SMPL
L1887-05MS - 1	CL-GWMS	13177	3.7251	-7470A	PPB	PPB	13 Mar 2020 10:40:55	U	SMPL
CCV24 - 1	CCV24	18823	5.3232	-7470A	PPB	PPB	13 Mar 2020 10:43:02	U	SMPL
CCB24 - 1	CCB24	-144	-0.0455	-7470A	PPB	PPB	13 Mar 2020 10:45:08	U	SMPL
L1887-05MSD - 1	CL-GWMSD	13187	3.7279	-7470A	PPB	PPB	13 Mar 2020 10:47:14	U	SMPL
PB127486TB - 1	PB127465TB	-44	-0.0172	-7470A	PPB	PPB	13 Mar 2020 10:49:21	U	SMPL
PB127515TB - 1	PB127504TB	-4	-0.0059	-7470A	PPB	PPB	13 Mar 2020 10:51:27	U	SMPL
L1752-04LX5 - 1		-23	-0.0112	-7470A	PPB	PPB	13 Mar 2020 10:53:33	U	SMPL
L1752-04A - 1		12390	3.5023	-7470A	PPB	PPB	13 Mar 2020 10:55:40	U	SMPL
L1887-04LX5 - 1		258	0.0683	-7470A	PPB	PPB	13 Mar 2020 10:57:46	U	SMPL
L1887-04A - 1		16358	4.6255	-7470A	PPB	PPB	13 Mar 2020 10:59:54	U	SMPL
L1887-05LX5 - 1		-88	-0.0296	-7470A	PPB	PPB	13 Mar 2020 11:02:01	U	SMPL
L1887-05A - 1		13306	3.7616	-7470A	PPB	PPB	13 Mar 2020 11:04:10	U	SMPL
CCV25 - 1	CCV25	18280	5.1695	-7470A	PPB	PPB	13 Mar 2020 11:06:17	U	SMPL
CCB25 - 1	CCB25	-129	-0.0413	-7470A	PPB	PPB	13 Mar 2020 11:08:23	U	SMPL