



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108262

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/13/2020	13:15
2	CAL2	1	Water	NA	NA	20.1	7.00	03/13/2020	13:16
3	CAL3	1	Water	NA	NA	20.2	10.03	03/13/2020	13:19
4	ICV	1	Water	NA	NA	20.1	6.99	03/13/2020	13:20
5	CCV1	1	Water	NA	NA	20.1	2.01	03/13/2020	13:22
6	L1917-01	1	Water	NA	NA	19.1	7.20	03/13/2020	13:24
7	L1917-01DUP	1	Water	NA	NA	19.2	7.21	03/13/2020	13:25
8	L1917-02	1	Water	NA	NA	18.9	7.08	03/13/2020	13:26
9	L1917-03	1	Water	NA	NA	19.1	7.02	03/13/2020	13:29
10	L1917-04	1	Water	NA	NA	19.8	7.10	03/13/2020	13:30
11	L1917-05	1	Water	NA	NA	19.6	7.31	03/13/2020	13:32
12	L1917-06	1	Water	NA	NA	20.1	7.12	03/13/2020	13:33
13	L1917-07	1	Water	NA	NA	20.0	7.31	03/13/2020	13:35
14	L1917-08	1	Water	NA	NA	20.2	8.16	03/13/2020	13:36
15	L1917-09	1	Water	NA	NA	20.3	7.74	03/13/2020	13:38
16	CCV2	1	Water	NA	NA	20.1	12.02	03/13/2020	13:40
17	L1917-10	1	Water	NA	NA	20.2	8.00	03/13/2020	13:42
18	CCV3	1	Water	NA	NA	20.1	2.01	03/13/2020	13:45

2B108262

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1917 WorkList ID : 137021 Department : Wet-Chemistry Date : 03-13-2020 13:06:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/13/2020	Water	L1917-01 -D	pH	Cool 4 deg C	TULL02	D62	MH-5	03/13/2020	9040C
03/13/2020	Water	L1917-02	pH	Cool 4 deg C	TULL02	D62	MH-725	03/13/2020	9040C
03/13/2020	Water	L1917-03	pH	Cool 4 deg C	TULL02	D62	MH-719	03/13/2020	9040C
03/13/2020	Water	L1917-04	pH	Cool 4 deg C	TULL02	D62	MH-720	03/13/2020	9040C
03/13/2020	Water	L1917-05	pH	Cool 4 deg C	TULL02	D62	CBH-594	03/13/2020	9040C
03/13/2020	Water	L1917-06	pH	Cool 4 deg C	TULL02	D62	CBH-140	03/13/2020	9040C
03/13/2020	Water	L1917-07	pH	Cool 4 deg C	TULL02	D62	CBH-139	03/13/2020	9040C
03/13/2020	Water	L1917-08 -E	pH	Cool 4 deg C	TULL02	D62	MH-F	03/13/2020	9040C
03/13/2020	Water	L1917-09 -E	pH	Cool 4 deg C	TULL02	D62	MH-G	03/13/2020	9040C
03/13/2020	Water	L1917-10	pH	Cool 4 deg C	TULL02	D62	MH-J	03/13/2020	9040C

Date/Time 03-13-2020 13:10
 Received by: JB
 Relinquished by: AM

Date/Time 03-13-2020 16:20
 Received by: AM
 Relinquished by: JP