



Analytical Summary Report

Analysis Method: 9041A

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108273

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/13/2020	16:25
2	CAL2	1	Water	NA	NA	20.1	7.01	03/13/2020	16:26
3	CAL3	1	Water	NA	NA	20.1	10.03	03/13/2020	16:27
4	ICV	1	Water	NA	NA	20.1	7.00	03/13/2020	16:30
5	CCV1	1	Water	NA	NA	20.1	2.01	03/13/2020	16:32
6	L1942-09	1	Water	NA	NA	17.8	6.95	03/13/2020	16:33
7	L1942-09DUP	1	Water	NA	NA	17.7	6.95	03/13/2020	16:34
8	CCV2	1	Water	NA	NA	20.2	12.03	03/13/2020	16:35

VB 108273

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1942 WorkList ID : 137029 Department : Wet-Chemistry Date : 03-13-2020 16:18:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/13/2020	Water	L1942-09	pH ☒	Cool 4 deg C	PSEG01		GW-TP-10	03/13/2020	9041A

Date/Time 03-13-2020 16:10
Received by: JD
Relinquished by: [Signature]

Date/Time 03-13-2020 17:15
Received by: [Signature]
Relinquished by: Jr