



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB108287

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	03/16/2020	09:50
2	L1942-10		1	8	100.00	>212.0	>212.0	03/16/2020	10:20
3	L1942-10DUP		1	8	100.00	>212.0	>212.0	03/16/2020	10:45

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

LB108287

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp31620 WorkList ID : 137069 Department : Wet-Chemistry Date : 03-16-2020 09:18:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/20/2020	Water	L1942-10	Flash Point	Cool 4 deg C	PSEG01		GW-TP-10	03/13/2020	1010A

Date/Time 03/16/2020 09:25
 Received by: KS
 Relinquished by: DK

Date/Time 03/16/2020 10:50
 Received by: DK
 Relinquished by: KS