


LB108288

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.022	0.022	23	0.000	23				
0.2	0.200	0.183	-0.017	584	0.0 %	584				
2.5	2.500	2.481	-0.019	8586	0.0 %	8586				
5.0	5.000	4.958	-0.042	17207	0.0 %	17207				
7.5	7.500	7.630	0.130	26510	0.0 %	26510				
10.0	10.000	9.926	-0.074	34500	0.0 %	34500				

LB108288 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Mar 2020 11:10:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	23	-	0.0000	7470A	PPB	16 Mar 2020 11:22:02	S	Std
0.2 - 1	51.2	584	-	0.2000	7470A	PPB	16 Mar 2020 11:24:10	S	Std
2.5 - 1	50.5	8586	-	2.5000	7470A	PPB	16 Mar 2020 11:26:17	S	Std
5.0 - 1	55	17207	-	5.0000	7470A	PPB	16 Mar 2020 11:28:25	S	Std
7.5 - 1	57.5	26510	-	7.5000	7470A	PPB	16 Mar 2020 11:30:32	S	Std
10.0 - 1	511	34500	-	10.0000	7470A	PPB	16 Mar 2020 11:35:05	S	Std
ICV59 - 1	ICV59	13440	3.8759	-	7470A	PPB	16 Mar 2020 11:38:12	U	SMPL
ICB59 - 1	ICB59	-254	-0.0579	-	7470A	PPB	16 Mar 2020 11:40:17	U	SMPL
CCV31 - 1	CCV31	16950	4.8842	-	7470A	PPB	16 Mar 2020 11:42:23	U	SMPL
CCB31 - 1	CCB31	-245	-0.0553	-	7470A	PPB	16 Mar 2020 11:44:28	U	SMPL
CRA - 1	CRA	535	0.1687	-	7470A	PPB	16 Mar 2020 11:46:35	U	SMPL
HighStd - 1	HighStd	34904	10.0417	-	7470A	PPB	16 Mar 2020 11:48:41	U	SMPL
ChkStd - 1	ChkStd	25798	7.4259	-	7470A	PPB	16 Mar 2020 11:50:46	U	SMPL
PB127529BL - 1	PBW	-326	-0.0786	-	7470A	PPB	16 Mar 2020 11:52:55	U	SMPL
PB127529BS - 1	LCSW	13509	3.8957	-	7470A	PPB	16 Mar 2020 11:55:03	U	SMPL
L1761-10 - 1	SU-01-031220	-242	-0.0545	-	7470A	PPB	16 Mar 2020 11:57:10	U	SMPL
L1761-10DUP - 1	SU-01-031220DUP	-68	-0.0045	-	7470A	PPB	16 Mar 2020 11:59:18	U	SMPL
L1761-10MS - 1	SU-01-031220MS	12476	3.5989	-	7470A	PPB	16 Mar 2020 12:01:25	U	SMPL
L1761-10MSD - 1	SU-01-031220MSD	14106	4.0672	-	7470A	PPB	16 Mar 2020 12:03:33	U	SMPL
L1762-15 - 1	TR-04-031220	-204	-0.0435	-	7470A	PPB	16 Mar 2020 12:05:39	U	SMPL
CCV32 - 1	CCV32	17709	5.1022	-	7470A	PPB	16 Mar 2020 12:07:45	U	SMPL
CCB32 - 1	CCB32	-247	-0.0559	-	7470A	PPB	16 Mar 2020 12:09:51	U	SMPL
L1902-02 - 1	VNJ-229-13	-37	0.0044	-	7470A	PPB	16 Mar 2020 12:11:58	U	SMPL
L1907-04 - 1	TP-6-SA	3	0.0159	-	7470A	PPB	16 Mar 2020 12:14:04	U	SMPL
L1907-08 - 1	TP-9-SA	-18	0.0099	-	7470A	PPB	16 Mar 2020 12:16:10	U	SMPL
L1912-02 - 1	Z1-1	-32	0.0059	-	7470A	PPB	16 Mar 2020 12:18:16	U	SMPL
L1912-06 - 1	Z3-S1	22	0.0214	-	7470A	PPB	16 Mar 2020 12:20:23	U	SMPL
L1912-10 - 1	Z3-S2	28	0.0231	-	7470A	PPB	16 Mar 2020 12:22:30	U	SMPL
L1912-14 - 1	Z1-2	21	0.0211	-	7470A	PPB	16 Mar 2020 12:24:37	U	SMPL
L1912-18 - 1	Z1-3	43	0.0274	-	7470A	PPB	16 Mar 2020 12:26:45	U	SMPL
L1912-22 - 1	Z1-4	47	0.0286	-	7470A	PPB	16 Mar 2020 12:28:52	U	SMPL
L1912-26 - 1	Z1-5	50	0.0294	-	7470A	PPB	16 Mar 2020 12:30:58	U	SMPL
CCV33 - 1	CCV33	17566	5.0611	-	7470A	PPB	16 Mar 2020 12:33:05	U	SMPL
CCB33 - 1	CCB33	-183	-0.0375	-	7470A	PPB	16 Mar 2020 12:35:11	U	SMPL
L1912-30 - 1	Z1-6	-25	0.0079	-	7470A	PPB	16 Mar 2020 12:37:18	U	SMPL
PB127529TB - 1	PB127505TB	85	0.0395	-	7470A	PPB	16 Mar 2020 12:39:24	U	SMPL
L1761-10LX5 - 1		-43	0.0027	-	7470A	PPB	16 Mar 2020 12:41:30	U	SMPL
L1761-10A - 1		13837	3.9899	-	7470A	PPB	16 Mar 2020 12:43:36	U	SMPL
CCV34 - 1	CCV34	17721	5.1056	-	7470A	PPB	16 Mar 2020 12:45:43	U	SMPL
CCB34 - 1	CCB34	-189	-0.0392	-	7470A	PPB	16 Mar 2020 12:47:50	U	SMPL