

LB 108292


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53195



A= 0.0000e+000

B= 1.8851e-004

C= -1.9829e-002

Rho= 0.9998609

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	202	0.000	202				
0.2	0.200	0.258	0.058	1472	0.0 %	1472				
2.5	2.500	2.450	-0.050	13103	0.0 %	13103				
5.0	5.000	4.894	-0.106	26064	0.0 %	26064				
7.5	7.500	7.572	0.072	40273	0.0 %	40273				
10.0	10.000	10.008	0.008	53195	0.0 %	53195				

LB108292 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Mar 2020 13:52:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	202	-	0.0000	7471B	PPB	16 Mar 2020 13:55:45	S	Std
0.2 - 1	50.2	1472	-	0.2000	7471B	PPB	16 Mar 2020 13:58:02	S	Std
2.5 - 1	52.5	13103	-	2.5000	7471B	PPB	16 Mar 2020 14:00:19	S	Std
5.0 - 1	55	26064	-	5.0000	7471B	PPB	16 Mar 2020 14:02:36	S	Std
7.5 - 1	57.5	40273	-	7.5000	7471B	PPB	16 Mar 2020 14:04:52	S	Std
10.0 - 1	510	53195	-	10.0000	7471B	PPB	16 Mar 2020 14:07:10	S	Std
ICV60 - 1	ICV60	20696	3.8816	-	7471B	PPB	16 Mar 2020 14:09:58	U	SMPL
ICB60 - 1	ICB60	91	-0.0027	-	7471B	PPB	16 Mar 2020 14:12:14	U	SMPL
CCV35 - 1	CCV35	26125	4.9051	-	7471B	PPB	16 Mar 2020 14:14:30	U	SMPL
CCB35 - 1	CCB35	65	-0.0076	-	7471B	PPB	16 Mar 2020 14:16:44	U	SMPL
CRA - 1	CRA	1397	0.2435	-	7471B	PPB	16 Mar 2020 14:19:01	U	SMPL
HighStd - 1	HighStd	53027	9.9764	-	7471B	PPB	16 Mar 2020 14:21:15	U	SMPL
ChkStd - 1	ChkStd	36835	6.9240	-	7471B	PPB	16 Mar 2020 14:23:31	U	SMPL
PB127549BL - 1	PBS	-353	-0.0864	-	7471B	PPB	16 Mar 2020 14:25:50	U	SMPL
PB127549BS - 1	LCSS	19230	3.6053	-	7471B	PPB	16 Mar 2020 14:28:07	U	SMPL
L1920-01 - 1	F-1	4271	0.7853	-	7471B	PPB	16 Mar 2020 14:30:24	U	SMPL
L1920-02 - 1	F-2	997	0.1681	-	7471B	PPB	16 Mar 2020 14:32:41	U	SMPL
L1933-01 - 1	72-12002	1925	0.3431	-	7471B	PPB	16 Mar 2020 14:34:58	U	SMPL
L1933-01DUP - 1	72-12002DUP	1651	0.2914	-	7471B	PPB	16 Mar 2020 14:37:16	U	SMPL
L1933-01MS - 1	72-12002MS	21715	4.0737	-	7471B	PPB	16 Mar 2020 14:39:31	U	SMPL
CCV36 - 1	CCV36	26588	4.9924	-	7471B	PPB	16 Mar 2020 14:41:46	U	SMPL
CCB36 - 1	CCB36	237	0.0248	-	7471B	PPB	16 Mar 2020 14:44:35	U	SMPL
L1933-01MSD - 1	72-12002MSD	21920	4.1124	-	7471B	PPB	16 Mar 2020 14:46:50	U	SMPL
L1933-04 - 1	72-11976	1820	0.3233	-	7471B	PPB	16 Mar 2020 14:49:06	U	SMPL
L1937-03 - 1	TP-E-S	1481	0.2594	-	7471B	PPB	16 Mar 2020 14:51:22	U	SMPL
L1942-02 - 1	TP-10-SA	10983	2.0506	-	7471B	PPB	16 Mar 2020 14:53:37	U	SMPL
L1942-06 - 1	TP-10-DA	2684	0.4861	-	7471B	PPB	16 Mar 2020 14:55:53	U	SMPL
L1942-13 - 1	TP-3-SA	9891	1.8448	-	7471B	PPB	16 Mar 2020 14:58:09	U	SMPL
L1942-17 - 1	TP-3-DA	1197	0.2058	-	7471B	PPB	16 Mar 2020 15:00:25	U	SMPL
L1942-21 - 1	TP-5-SA	1471	0.2575	-	7471B	PPB	16 Mar 2020 15:02:42	U	SMPL
L1942-25 - 1	TP-5-DA	270	0.0311	-	7471B	PPB	16 Mar 2020 15:04:59	U	SMPL
L1943-02 - 1	TP-8A	1030	0.1743	-	7471B	PPB	16 Mar 2020 15:07:14	U	SMPL
CCV37 - 1	CCV37	26973	5.0649	-	7471B	PPB	16 Mar 2020 15:09:29	U	SMPL
CCB37 - 1	CCB37	80	-0.0047	-	7471B	PPB	16 Mar 2020 15:11:45	U	SMPL
L1943-06 - 1	TP-9A	2053	0.3672	-	7471B	PPB	16 Mar 2020 15:14:01	U	SMPL
L1943-10 - 1	J-B-2A	7488	1.3918	-	7471B	PPB	16 Mar 2020 15:16:16	U	SMPL
L1933-01LX5 - 1		496	0.0737	-	7471B	PPB	16 Mar 2020 15:18:32	U	SMPL
L1933-01A - 1		23935	4.4922	-	7471B	PPB	16 Mar 2020 15:25:57	U	SMPL
CCV38 - 1	CCV38	26701	5.0137	-	7471B	PPB	16 Mar 2020 15:28:13	U	SMPL
CCB38 - 1	CCB38	186	0.0152	-	7471B	PPB	16 Mar 2020 15:30:30	U	SMPL