



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108297

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/16/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	03/16/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	03/16/2020	17:12
4	ICV	1	Water	NA	NA	20.2	7.00	03/16/2020	17:15
5	CCV1	1	Water	NA	NA	20.2	2.01	03/16/2020	17:40
6	L1946-01	1	Water	NA	NA	17.2	7.14	03/16/2020	17:42
7	L1946-02	1	Water	NA	NA	17.2	6.87	03/16/2020	17:45
8	L1946-03	1	Water	NA	NA	17.0	6.90	03/16/2020	17:48
9	L1946-03DUP	1	Water	NA	NA	17.1	6.91	03/16/2020	17:49
10	CCV2	1	Water	NA	NA	20.2	12.02	03/16/2020	17:50

VB108247

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1946 WorkList ID : 137085 Department : Wet-Chemistry Date : 03-16-2020 16:08:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/16/2020	Water	L1946-01 ☒	pH	Cool 4 deg C	PSEG01	I63	40748	03/16/2020	9040C
03/16/2020	Water	L1946-02 ☒	pH	Cool 4 deg C	PSEG01	I63	40747	03/16/2020	9040C
03/16/2020	Water	L1946-03 ☒	pH	Cool 4 deg C	PSEG01	I63	40749	03/16/2020	9040C

Date/Time 03-16-2020 16:10
Received by: JA
Relinquished by: JA

Date/Time 03-16-2020 18:00
Received by: JA
Relinquished by: JA