



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB108311

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	03/17/2020	13:40
2	L1948-12		1	8	100.00	>212.0	>212.0	03/17/2020	14:10
3	L1948-12DUP		1	8	100.00	>212.0	>212.0	03/17/2020	14:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 108311

WorkList Name : fp31720-1 WorkList ID : 137114 Department : Wet-Chemistry Date : 03-17-2020 13:12:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/20/2020	Solid	L1948-12	Flash Point	Cool 4 deg C	PSEG01	I63	OIL	03/16/2020	1010A

Date/Time 03/17/2020 13:15
 Received by: KS
 Relinquished by: AP

Date/Time 03/17/2020 14:40
 Received by: AP
 Relinquished by: KS