



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108319

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/17/2020	17:00
2	CAL2	1	Water	NA	NA	20.1	7.00	03/17/2020	17:01
3	CAL3	1	Water	NA	NA	20.1	10.03	03/17/2020	17:02
4	ICV	1	Water	NA	NA	20.2	6.99	03/17/2020	17:05
5	CCV1	1	Water	NA	NA	20.1	2.01	03/17/2020	17:06
6	L1954-04	1	Solid	20.03	20	20.6	9.23	03/17/2020	17:10
7	L1954-04DUP	1	Solid	20.04	20	20.7	9.24	03/17/2020	17:11
8	L1954-08	1	Solid	20.02	20	20.7	8.06	03/17/2020	17:15
9	CCV2	1	Water	NA	NA	20.2	12.02	03/17/2020	17:20

WORKLIST(Hardcopy Internal Chain)

282108319

WorkList Name : CORROSIVITY L1953

WorkList ID : 137121

Department : Wet-Chemistry

Date : 03-17-2020 15:19:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/16/2020	Solid	L1954-04	Corrosivity	Cool 4 deg C	PSEG01	I52	TP-10A	03/16/2020	9045D
03/16/2020	Solid	L1954-08	Corrosivity	Cool 4 deg C	PSEG01	I52	TP-11A	03/16/2020	9045D

Date/Time 03-17-2020 15:120

Received by: J9

Relinquished by: J9

Date/Time 03-17-2020 17:130

Received by: J9

Relinquished by: J9