



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108323

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	03/18/2020	09:20
2	CAL2	1	Water	NA	NA	20.1	7.01	03/18/2020	09:21
3	CAL3	1	Water	NA	NA	20.2	10.03	03/18/2020	09:22
4	ICV	1	Water	NA	NA	20.1	7.00	03/18/2020	09:24
5	CCV1	1	Water	NA	NA	20.1	2.01	03/18/2020	09:25
6	L1960-01	1	Water	NA	NA	17.1	6.85	03/18/2020	09:30
7	L1960-01DUP	1	Water	NA	NA	17.2	6.86	03/18/2020	09:31
8	CCV2	1	Water	NA	NA	20.2	12.02	03/18/2020	09:33

VB 108323

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1960 WorkList ID : 137146 Department : Wet-Chemistry Date : 03-18-2020 09:14:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/17/2020	Water	L1960-01	pH	Cool 4 deg C	APT101	E11	WGGM-20-WW-TSF-010	03/17/2020	SM 4500-H B

Date/Time 03-18-2020 9:15
 Received by: JR
 Relinquished by: JR

Date/Time 03-18-2020 10:20
 Received by: JR
 Relinquished by: JR