

LB102328


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

52977



A= 0.0000e+000

B= 1.8861e-004

C= 3.2405e-002

Rho= 0.9998913

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.036	0.036	20	0.000	20				
0.2	0.200	0.237	0.037	1084	0.0 %	1084				
2.5	2.500	2.459	-0.041	12867	0.0 %	12867				
5.0	5.000	4.897	-0.103	25791	0.0 %	25791				
7.5	7.500	7.547	0.047	39841	0.0 %	39841				
10.0	10.000	10.024	0.024	52977	0.0 %	52977				

LB108328 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Mar 2020 13:28:42

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	20	-	0.0000	7471B	PPB	18 Mar 2020 13:30:59	S	Std
0.2 - 1	50.2	1084	-	0.2000	7471B	PPB	18 Mar 2020 13:33:16	S	Std
2.5 - 1	52.5	12867	-	2.5000	7471B	PPB	18 Mar 2020 13:35:33	S	Std
5.0 - 1	55	25791	-	5.0000	7471B	PPB	18 Mar 2020 13:37:49	S	Std
7.5 - 1	57.5	39841	-	7.5000	7471B	PPB	18 Mar 2020 13:40:06	S	Std
10.0 - 1	510	52977	-	10.0000	7471B	PPB	18 Mar 2020 13:45:00	S	Std
ICV64 - 1	ICV64	20916	3.9773	-	7471B	PPB	18 Mar 2020 13:47:44	U	SMPL
ICB64 - 1	ICB64	-182	-0.0019	-	7471B	PPB	18 Mar 2020 13:50:00	U	SMPL
CCV55 - 1	CCV55	26037	4.9432	-	7471B	PPB	18 Mar 2020 13:52:16	U	SMPL
CCB55 - 1	CCB55	-192	-0.0038	-	7471B	PPB	18 Mar 2020 13:54:31	U	SMPL
CRA - 1	CRA	1050	0.2304	-	7471B	PPB	18 Mar 2020 13:56:47	U	SMPL
HighStd - 1	HighStd	53691	10.1589	-	7471B	PPB	18 Mar 2020 13:59:02	U	SMPL
ChkStd - 1	ChkStd	36939	6.9994	-	7471B	PPB	18 Mar 2020 14:01:18	U	SMPL
PB127625BL - 1	PBS	-417	-0.0462	-	7471B	PPB	18 Mar 2020 14:03:36	U	SMPL
PB127625BS - 1	LCSS	18975	3.6112	-	7471B	PPB	18 Mar 2020 14:05:54	U	SMPL
L1954-02 - 1	TP-10A	772	0.1780	-	7471B	PPB	18 Mar 2020 14:08:10	U	SMPL
L1954-02DUP - 1	TP-10ADUP	689	0.1624	-	7471B	PPB	18 Mar 2020 14:13:03	U	SMPL
L1954-02MS - 1	TP-10AMS	21035	3.9998	-	7471B	PPB	18 Mar 2020 14:15:20	U	SMPL
L1954-02MSD - 1	TP-10AMSD	22482	4.2727	-	7471B	PPB	18 Mar 2020 14:17:36	U	SMPL
L1954-06 - 1	TP-11A	7432	1.4341	-	7471B	PPB	18 Mar 2020 14:19:52	U	SMPL
CCV56 - 1	CCV56	26281	4.9892	-	7471B	PPB	18 Mar 2020 14:22:07	U	SMPL
CCB56 - 1	CCB56	-170	0.0003	-	7471B	PPB	18 Mar 2020 14:24:23	U	SMPL
L1957-01 - 1	DRILL-CUTTINGS	276	0.0845	-	7471B	PPB	18 Mar 2020 14:26:39	U	SMPL
L1954-02LX5 - 1		84	0.0482	-	7471B	PPB	18 Mar 2020 14:28:54	U	SMPL
L1954-02A - 1		21870	4.1572	-	7471B	PPB	18 Mar 2020 14:31:10	U	SMPL
CCV57 - 1	CCV57	26391	5.0099	-	7471B	PPB	18 Mar 2020 14:33:25	U	SMPL
CCB57 - 1	CCB57	-160	0.0022	-	7471B	PPB	18 Mar 2020 14:35:42	U	SMPL