


LB108339

7470A

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.043	0.043	0	0.000	0				
0.2	0.200	0.235	0.035	713	0.0 %	713				
2.5	2.500	2.406	-0.094	8789	0.0 %	8789				
5.0	5.000	5.038	0.038	18574	0.0 %	18574				
7.5	7.500	7.408	-0.092	27388	0.0 %	27388				
10.0	10.000	10.070	0.070	37284	0.0 %	37284				

LB108339 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Mar 2020 11:06:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	0	-	0.0000	7470A	PPB	19 Mar 2020 11:12:56	S	Std
0.2 - 1	50.2	713	-	0.2000	7470A	PPB	19 Mar 2020 11:15:04	S	Std
2.5 - 1	52.5	8789	-	2.5000	7470A	PPB	19 Mar 2020 11:17:12	S	Std
5.0 - 1	55	18574	-	5.0000	7470A	PPB	19 Mar 2020 11:19:19	S	Std
7.5 - 1	57.5	27388	-	7.5000	7470A	PPB	19 Mar 2020 11:21:26	S	Std
10.0 - 1	510	37284	-	10.0000	7470A	PPB	19 Mar 2020 11:23:32	S	Std
ICV65 - 1	ICV65	13296	3.6186	-7470A	PPB	19 Mar 2020 11:38:36	U	SMPL	
ICB65 - 1	ICB65	-186	-0.0072	-7470A	PPB	19 Mar 2020 11:40:41	U	SMPL	
CCV58 - 1	CCV58	19097	5.1787	-7470A	PPB	19 Mar 2020 11:42:46	U	SMPL	
CCB58 - 1	CCB58	-234	-0.0201	-7470A	PPB	19 Mar 2020 11:44:52	U	SMPL	
CRA - 1	CRA	668	0.2225	-7470A	PPB	19 Mar 2020 11:46:59	U	SMPL	
HighStd - 1	HighStd	37225	10.0539	-7470A	PPB	19 Mar 2020 11:49:05	U	SMPL	
ChkStd - 1	ChkStd	24489	6.6287	-7470A	PPB	19 Mar 2020 11:51:11	U	SMPL	
PB127651BL - 1	PBW	-311	-0.0408	-7470A	PPB	19 Mar 2020 11:53:19	U	SMPL	
PB127651BS - 1	LCSW	12687	3.4548	-7470A	PPB	19 Mar 2020 11:55:28	U	SMPL	
L1953-06 - 1	DRUM-COMP	-194	-0.0094	-7470A	PPB	19 Mar 2020 11:57:34	U	SMPL	
L1954-04 - 1	TP-10A	-48	0.0299	-7470A	PPB	19 Mar 2020 11:59:42	U	SMPL	
L1954-08 - 1	TP-11A	-24	0.0364	-7470A	PPB	19 Mar 2020 12:01:50	U	SMPL	
L1954-08DUP - 1	TP-11ADUP	2	0.0434	-7470A	PPB	19 Mar 2020 12:03:58	U	SMPL	
L1954-08MS - 1	TP-11AMS	13237	3.6027	-7470A	PPB	19 Mar 2020 12:06:04	U	SMPL	
CCV59 - 1	CCV59	19162	5.1961	-7470A	PPB	19 Mar 2020 12:08:10	U	SMPL	
CCB59 - 1	CCB59	-268	-0.0293	-7470A	PPB	19 Mar 2020 12:10:17	U	SMPL	
L1954-08MSD - 1	TP-11AMSD	12769	3.4768	-7470A	PPB	19 Mar 2020 12:12:24	U	SMPL	
PB127651TB - 1	PB127623TB	-2	0.0423	-7470A	PPB	19 Mar 2020 12:14:30	U	SMPL	
L1954-08LX5 - 1		-79	0.0216	-7470A	PPB	19 Mar 2020 12:16:37	U	SMPL	
L1954-08A - 1		13587	3.6968	-7470A	PPB	19 Mar 2020 12:18:44	U	SMPL	
CCV60 - 1	CCV60	19273	5.2260	-7470A	PPB	19 Mar 2020 12:20:51	U	SMPL	
CCB60 - 1	CCB60	-247	-0.0236	-7470A	PPB	19 Mar 2020 12:22:57	U	SMPL	