



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108341

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/18/2020	16:35
2	CAL2	1	Water	NA	NA	20.1	7.01	03/18/2020	16:36
3	CAL3	1	Water	NA	NA	20.1	10.03	03/18/2020	16:37
4	ICV	1	Water	NA	NA	20.1	7.00	03/18/2020	16:40
5	CCV1	1	Water	NA	NA	20.2	2.01	03/18/2020	16:42
6	L1966-02	1	Water	NA	NA	15.2	4.53	03/18/2020	16:44
7	L1966-02DUP	1	Water	NA	NA	15.3	4.54	03/18/2020	16:45
8	CCV2	1	Water	NA	NA	20.1	12.02	03/18/2020	16:46

WORKLIST(Hardcopy Internal Chain)

W3 108341

WorkList Name : CORROSIONITY L1962 W WorkList ID : 137170 Department : Wet-Chemistry Date : 03-18-2020 16:28:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/18/2020	Water	L1966-02	-1) Corrosivity	Cool 4 deg C	PSEG01	I63	TEST-PIT-3	03/18/2020	9040C

Date/Time 03.18.2020 16:30
 Received by: IT
 Relinquished by: _____

Date/Time 03.18.2020 17:30
 Received by: IT
 Relinquished by: JD