



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108344

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	03/19/2020	11:10
2	CAL2	1	Water	NA	NA	20.1	7.00	03/19/2020	11:11
3	CAL3	1	Water	NA	NA	20.1	10.03	03/19/2020	11:12
4	ICV	1	Water	NA	NA	20.1	7.01	03/19/2020	11:15
5	CCV1	1	Water	NA	NA	20.2	2.01	03/19/2020	11:16
6	L1982-01	1	Water	NA	NA	20.4	5.37	03/19/2020	11:20
7	L1982-01DUP	1	Water	NA	NA	20.5	5.38	03/19/2020	11:22
8	L1982-02	1	Water	NA	NA	20.6	5.42	03/19/2020	11:25
9	L1982-03	1	Water	NA	NA	20.5	5.49	03/19/2020	11:30
10	CCV2	1	Water	NA	NA	20.2	12.02	03/19/2020	11:32

203108346

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1982 WorkList ID : 137193 Department : Wet-Chemistry Date : 03-19-2020 11:00:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/17/2020	Water	L1982-01	pH	Cool 4 deg C	STER07	E21	CELLS-A	03/17/2020	SM 4500-H B
03/17/2020	Water	L1982-02	pH	Cool 4 deg C	STER07	E21	CELLS-B/C	03/17/2020	SM 4500-H B
03/17/2020	Water	L1982-03	pH	Cool 4 deg C	STER07	E21	CELLS-D	03/17/2020	SM 4500-H B

Date/Time 03-19-2020 11:05
 Received by: JD
 Relinquished by: JD

Date/Time 03-19-2020 12:30
 Received by: JA
 Relinquished by: JD