



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB108364

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	03/19/2020	11:20
2	L1966-02		1	8	100.00	>212.0	>212.0	03/19/2020	11:45
3	L1966-02DUP		1	8	100.00	>212.0	>212.0	03/19/2020	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 108364

WorkList Name : fp31920 WorkList ID : 137187 Department : Wet-Chemistry Date : 03-19-2020 10:41:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/31/2020	Water	L1966-02 C	Flash Point	Cool 4 deg C	PSEG01	I63	TEST-PIT-3	03/18/2020	1010A

Date/Time 03/19/2020 10:45
 Received by: KS
 Relinquished by: 70

Date/Time 03/19/2020 12:15
 Received by: 70
 Relinquished by: KS