



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108365

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/19/2020	17:00
2	CAL2	1	Water	NA	NA	20.2	7.00	03/19/2020	17:01
3	CAL3	1	Water	NA	NA	20.2	10.03	03/19/2020	17:02
4	ICV	1	Water	NA	NA	20.2	7.01	03/19/2020	17:05
5	CCV1	1	Water	NA	NA	20.1	2.01	03/19/2020	17:06
6	L1749-03	1	Solid	20.03	20	20.6	6.64	03/19/2020	17:11
7	L1749-03DUP	1	Solid	20.04	20	20.7	6.65	03/19/2020	17:12
8	L1755-08	1	Solid	20.02	20	20.4	8.15	03/19/2020	17:15
9	L1762-20	1	Solid	20.04	20	20.4	8.63	03/19/2020	17:20
10	CCV2	1	Water	NA	NA	20.1	12.02	03/19/2020	17:22

HB108355

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L1755 WorkList ID : 137210 Department : Wet-Chemistry Date : 03-19-2020 14:23:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/19/2020	Solid	L1749-03	Corrosivity	Cool 4 deg C	PSEG05	E22	BU-02-031920	03/19/2020	9045D
03/19/2020	Solid	L1755-08	Corrosivity	Cool 4 deg C	PSEG05	E41	OK-01-031920	03/19/2020	9045D
03/19/2020	Solid	L1762-20	Corrosivity	Cool 4 deg C	PSEG05		TR-05-031920	03/19/2020	9045D

Date/Time 03-14-2020 16:00
Received by: JQ
Relinquished by: KS

Date/Time 03-14-2020 18:00
Received by: KS
Relinquished by: JQ