

LB108371


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36511



A= 0.0000e+000

B= 2.7346e-004

C= 1.0564e-002

Rho= 0.9999701

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.025	0.025	53	0.000	53				
0.2	0.200	0.213	0.013	740	0.0 %	740				
2.5	2.500	2.453	-0.047	8933	0.0 %	8933				
5.0	5.000	4.977	-0.023	18162	0.0 %	18162				
7.5	7.500	7.537	0.037	27523	0.0 %	27523				
10.0	10.000	9.995	-0.005	36511	0.0 %	36511				

LB108371 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Mar 2020 08:47:16

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	53	-	0.0000	7470A	PPB	20 Mar 2020 08:51:31	S	Std
0.2 - 1	S0.2	740	-	0.2000	7470A	PPB	20 Mar 2020 08:53:38	S	Std
2.5 - 1	S2.5	8933	-	2.5000	7470A	PPB	20 Mar 2020 08:55:46	S	Std
5.0 - 1	S5	18162	-	5.0000	7470A	PPB	20 Mar 2020 08:57:53	S	Std
7.5 - 1	S7.5	27523	-	7.5000	7470A	PPB	20 Mar 2020 09:00:00	S	Std
10.0 - 1	S10	36511	-	10.0000	7470A	PPB	20 Mar 2020 09:02:07	S	Std
ICV67 - 1	ICV67	14048	3.8521	-	7470A	PPB	20 Mar 2020 09:04:45	U	SMPL
ICB67 - 1	ICB67	-186	-0.0403	-	7470A	PPB	20 Mar 2020 09:06:51	U	SMPL
CCV64 - 1	CCV64	18947	5.1917	-	7470A	PPB	20 Mar 2020 09:08:57	U	SMPL
CCB64 - 1	CCB64	-189	-0.0411	-	7470A	PPB	20 Mar 2020 09:11:03	U	SMPL
CRA - 1	CRA	654	0.1894	-	7470A	PPB	20 Mar 2020 09:13:10	U	SMPL
HighStd - 1	HighStd	36447	9.9772	-	7470A	PPB	20 Mar 2020 09:15:16	U	SMPL
ChkStd - 1	ChkStd	26348	7.2156	-	7470A	PPB	20 Mar 2020 09:17:22	U	SMPL
PB127685BL - 1	PBW	-322	-0.0775	-	7470A	PPB	20 Mar 2020 09:19:30	U	SMPL
PB127685BS - 1	LCSW	12855	3.5258	-	7470A	PPB	20 Mar 2020 09:21:38	U	SMPL
L1964-03 - 1	2000448388	-147	-0.0296	-	7470A	PPB	20 Mar 2020 09:23:45	U	SMPL
L1964-03DUP - 1	2000448388DUP	10	0.0133	-	7470A	PPB	20 Mar 2020 09:25:53	U	SMPL
L1964-03MS - 1	2000448388MS	13023	3.5718	-	7470A	PPB	20 Mar 2020 09:28:00	U	SMPL
L1964-03MSD - 1	2000448388MSD	13523	3.7085	-	7470A	PPB	20 Mar 2020 09:30:08	U	SMPL
L1965-03 - 1	80564952	-23	0.0043	-	7470A	PPB	20 Mar 2020 09:32:14	U	SMPL
CCV65 - 1	CCV65	18893	5.1770	-	7470A	PPB	20 Mar 2020 09:34:20	U	SMPL
CCB65 - 1	CCB65	-186	-0.0403	-	7470A	PPB	20 Mar 2020 09:36:26	U	SMPL
L1978-04 - 1	TP-5A	139	0.0486	-	7470A	PPB	20 Mar 2020 09:38:33	U	SMPL
L1978-08 - 1	TP-6A	151	0.0519	-	7470A	PPB	20 Mar 2020 09:40:39	U	SMPL
L1978-12 - 1	TP-7A	42	0.0220	-	7470A	PPB	20 Mar 2020 09:42:46	U	SMPL
L1980-01 - 1	BUS-WASH-SLUDGE-PIT	244	0.0773	-	7470A	PPB	20 Mar 2020 09:44:52	U	SMPL
L1980-02 - 1	CHASSIS-WASH-SLUDGE-PIT	286	0.0888	-	7470A	PPB	20 Mar 2020 09:46:58	U	SMPL
PB127686BL - 1	PBW	-21	0.0048	-	7470A	PPB	20 Mar 2020 09:49:05	U	SMPL
PB127686BS - 1	LCSW	13783	3.7796	-	7470A	PPB	20 Mar 2020 09:51:12	U	SMPL
L1966-02 - 1	TEST-PIT-3	-10	0.0078	-	7470A	PPB	20 Mar 2020 09:53:20	U	SMPL
L1966-02DUP - 1	TEST-PIT-3DUP	53	0.0251	-	7470A	PPB	20 Mar 2020 09:55:28	U	SMPL
L1966-02MS - 1	TEST-PIT-3MS	13433	3.6839	-	7470A	PPB	20 Mar 2020 09:57:34	U	SMPL
CCV66 - 1	CCV66	19188	5.2576	-	7470A	PPB	20 Mar 2020 09:59:40	U	SMPL
CCB66 - 1	CCB66	-86	-0.0130	-	7470A	PPB	20 Mar 2020 10:01:46	U	SMPL
L1966-02MSD - 1	TEST-PIT-3MSD	13473	3.6948	-	7470A	PPB	20 Mar 2020 10:03:52	U	SMPL
PB127687BL - 1	PBW	-98	-0.0162	-	7470A	PPB	20 Mar 2020 10:05:58	U	SMPL
PB127687BS - 1	LCSW	13401	3.6751	-	7470A	PPB	20 Mar 2020 10:08:04	U	SMPL
L1966-01 - 1	TEST-PIT-3	515	0.1514	-	7470A	PPB	20 Mar 2020 10:19:31	U	SMPL
L1966-01DUP - 1	TEST-PIT-3DUP	443	0.1317	-	7470A	PPB	20 Mar 2020 10:21:37	U	SMPL
L1966-01MS - 1	TEST-PIT-3MS	17771	4.8701	-	7470A	PPB	20 Mar 2020 10:23:44	U	SMPL
L1966-01MSD - 1	TEST-PIT-3MSD	16894	4.6303	-	7470A	PPB	20 Mar 2020 10:25:51	U	SMPL
L1966-04 - 1	TEST-PIT-3-FILTERED	47	0.0234	-	7470A	PPB	20 Mar 2020 10:27:59	U	SMPL
CCV67 - 1	CCV67	20008	5.4819	-	7470A	PPB	20 Mar 2020 10:30:07	U	SMPL
CCB67 - 1	CCB67	66	0.0286	-	7470A	PPB	20 Mar 2020 10:32:13	U	SMPL
L1964-03LX5 - 1		96	0.0368	-	7470A	PPB	20 Mar 2020 10:34:19	U	SMPL
L1964-03A - 1		14053	3.8534	-	7470A	PPB	20 Mar 2020 10:36:26	U	SMPL
L1966-02LX5 - 1		50	0.0242	-	7470A	PPB	20 Mar 2020 10:38:32	U	SMPL
L1966-02A - 1		14684	4.0260	-	7470A	PPB	20 Mar 2020 10:40:38	U	SMPL
L1966-01LX5 - 1		157	0.0535	-	7470A	PPB	20 Mar 2020 10:42:46	U	SMPL
L1966-01A - 1		17233	4.7230	-	7470A	PPB	20 Mar 2020 10:44:52	U	SMPL
PB127685TB - 1	PB127660TB	18	0.0155	-	7470A	PPB	20 Mar 2020 10:47:12	U	SMPL
PB127686TB - 1	PB127680TB	88	0.0346	-	7470A	PPB	20 Mar 2020 10:49:19	U	SMPL
CCV68 - 1	CCV68	19943	5.4641	-	7470A	PPB	20 Mar 2020 10:51:25	U	SMPL
CCB68 - 1	CCB68	71	0.0300	-	7470A	PPB	20 Mar 2020 10:53:33	U	SMPL
				-	7470A	PPB	20 Mar 2020 10:55:40	U	SMPL