



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108378

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/20/2020	16:40
2	CAL2	1	Water	NA	NA	20.1	7.00	03/20/2020	16:41
3	CAL3	1	Water	NA	NA	20.1	10.03	03/20/2020	16:42
4	ICV	1	Water	NA	NA	20.1	7.00	03/20/2020	16:45
5	CCV1	1	Water	NA	NA	20.1	2.01	03/20/2020	16:46
6	L1994-04	1	Solid	20.03	20	21.6	7.74	03/20/2020	16:47
7	L1994-08	1	Solid	20.04	20	21.7	8.25	03/20/2020	16:50
8	L2006-01	1	Solid	20.02	20	21.2	5.94	03/20/2020	17:00
9	L2009-10	1	Solid	20.03	20	21.3	7.10	03/20/2020	17:01
10	L2009-10DUP	1	Solid	20.04	20	21.4	7.11	03/20/2020	17:02
11	CCV2	1	Water	NA	NA	20.2	12.02	03/20/2020	17:05

28108378

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSION L1994 WorkList ID : 137238 Department : Wet-Chemistry Date : 03-20-2020 14:33:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/19/2020	Solid	L1994-04	Corrosivity	Cool 4 deg C	PSEG01	F11	TP-4A	03/19/2020	9045D
03/19/2020	Solid	L1994-08	Corrosivity	Cool 4 deg C	PSEG01	F11	TP-3A	03/19/2020	9045D
03/19/2020	Solid	L2006-01	Corrosivity	Cool 4 deg C	CLEA11	F11	OIL-WATER-SEPERATOR	03/19/2020	9045D
03/20/2020	Solid	L2009-10	Corrosivity	Cool 4 deg C	TULL02	F11	TRACKS-66-67	03/20/2020	9045D

Date/Time 03/20/2020 14:35
Received by: Jn
Relinquished by: Jn

Date/Time 03/20/2020 17:10
Received by: Jn
Relinquished by: Jn