



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108396

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/23/2020	10:15
2	CAL2	1	Water	NA	NA	20.1	7.01	03/23/2020	10:16
3	CAL3	1	Water	NA	NA	20.1	10.04	03/23/2020	10:17
4	ICV	1	Water	NA	NA	20.3	6.99	03/23/2020	10:18
5	CCV1	1	Water	NA	NA	20.1	2.01	03/23/2020	10:20
6	L2015-01	1	Water	NA	NA	18.2	7.77	03/23/2020	10:22
7	L2015-01DUP	1	Water	NA	NA	18.3	7.78	03/23/2020	10:23
8	CCV2	1	Water	NA	NA	20.2	12.02	03/23/2020	10:25

WORKLIST(Hardcopy Internal Chain)

28108396

WorkList Name : PH L2015 WorkList ID : 137271 Department : Wet-Chemistry Date : 03-23-2020 10:05:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/19/2020	Water	L2015-01 ~F	pH	Cool 4 deg C	MINO01	F11	011-OUTFALL	03/19/2020	9040C

Date/Time 03/23/2020 10:10
 Received by: Jo AM
 Relinquished by:

Date/Time 03/23/2020 11:00
 Received by: AM
 Relinquished by: Jo