



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108401

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/23/2020	12:40
2	CAL2	1	Water	NA	NA	20.1	7.00	03/23/2020	12:42
3	CAL3	1	Water	NA	NA	20.2	10.03	03/23/2020	12:45
4	ICV	1	Water	NA	NA	20.1	6.99	03/23/2020	12:46
5	CCV1	1	Water	NA	NA	20.32	2.01	03/23/2020	12:48
6	L2018-01	1	Water	NA	NA	12.1	5.74	03/23/2020	12:50
7	L2018-01DUP	1	Water	NA	NA	12.2	5.75	03/23/2020	12:51
8	L2018-02	1	Water	NA	NA	12.4	6.48	03/23/2020	13:00
9	L2018-03	1	Water	NA	NA	13.3	6.45	03/23/2020	13:02
10	L2018-04	1	Water	NA	NA	13.7	6.58	03/23/2020	13:05
11	CCV2	1	Water	NA	NA	20.1	12.02	03/23/2020	13:06

23108401

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2018

WorkList ID : 137276

Department : Wet-Chemistry

Date : 03-23-2020 12:30:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/20/2020	Water	L2018-01-11	pH	Cool 4 deg C	LOCK01		MW-1	03/20/2020	SM 4500-H B
03/20/2020	Water	L2018-02-11	pH	Cool 4 deg C	LOCK01		MW-3	03/20/2020	SM 4500-H B
03/20/2020	Water	L2018-03-11	pH	Cool 4 deg C	LOCK01		MW-4	03/20/2020	SM 4500-H B
03/20/2020	Water	L2018-04-11	pH	Cool 4 deg C	LOCK01		MW-2	03/20/2020	SM 4500-H B

Date/Time 03/23/2020 12:35
Received by: JN
Relinquished by: JN

Date/Time 03/23/2020 12:40
Received by: JN
Relinquished by: JN