



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108403

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/23/2020	16:00
2	CAL2	1	Water	NA	NA	20.1	7.00	03/23/2020	16:01
3	CAL3	1	Water	NA	NA	20.1	10.03	03/23/2020	16:03
4	ICV	1	Water	NA	NA	20.1	7.01	03/23/2020	16:05
5	CCV1	1	Water	NA	NA	20.2	2.02	03/23/2020	17:00
6	L2011-01	1	Solid	20.03	20	20.6	7.16	03/23/2020	17:02
7	L2011-02	1	Solid	20.04	20	20.8	7.35	03/23/2020	17:05
8	L2011-03	1	Solid	20.02	20	20.7	7.56	03/23/2020	17:09
9	L2011-03DUP	1	Solid	20.03	20	20.8	7.57	03/23/2020	17:10
10	CCV2	1	Water	NA	NA	20.1	12.02	03/23/2020	17:12

WORKLIST(Hardcopy Internal Chain)

108403

WorkList Name : CORROSION L2011

WorkList ID : 137283

Department : Wet-Chemistry

Date : 03-23-2020 15:43:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/20/2020	Solid	L2011-01	Corrosivity	Cool 4 deg C	PARS15	F22	GHD1-WC-S-BC-IBC	03/20/2020	9045D
03/20/2020	Solid	L2011-02	Corrosivity	Cool 4 deg C	PARS15	F22	GHD1-WC-S-SB-01	03/20/2020	9045D
03/20/2020	Solid	L2011-03	Corrosivity	Cool 4 deg C	PARS15	F22	GHD1-WC-S-SB-0203	03/20/2020	9045D

Date/Time 03/23/2020 15:40
 Received by: Jo
 Relinquished by: KS

Date/Time 03/23/2020 17:13
 Received by: KS
 Relinquished by: Jo