



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108408

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/23/2020	16:00
2	CAL2	1	Water	NA	NA	20.1	7.00	03/23/2020	16:01
3	CAL3	1	Water	NA	NA	20.1	10.03	03/23/2020	16:03
4	ICV	1	Water	NA	NA	20.1	7.01	03/23/2020	16:05
5	CCV1	1	Water	NA	NA	20.1	2.01	03/23/2020	16:06
6	L2011-04	1	Water	NA	NA	20.2	6.84	03/23/2020	16:09
7	L2021-01	1	Water	NA	NA	19.2	6.39	03/23/2020	16:10
8	L2021-04	1	Water	NA	NA	18.8	6.54	03/23/2020	16:12
9	L2021-04DUP	1	Water	NA	NA	18.9	6.55	03/23/2020	16:13
10	CCV2	1	Water	NA	NA	20.1	12.01	03/23/2020	16:15

WORKLIST(Hardcopy Internal Chain)

187 1084 08

WorkList Name : PH L2021

WorkList ID : 137284

Department : Wet-Chemistry

Date : 03-23-2020 15:44:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/20/2020	Water	L 2011-04	pH	Cool 4 deg C	PARS15	F22	GHD1-WC-C-01	03/20/2020	9040C
03/23/2020	Water	L 2021-01	pH	Cool 4 deg C	PSEG01	F22	MW-A	03/23/2020	9040C
03/23/2020	Water	L 2021-04	pH	Cool 4 deg C	PSEG01	F22	MW-B	03/23/2020	9040C

Date/Time 03/23/2020 15:50

Received by: JO

Relinquished by: D.P.

Date/Time 03/23/2020 17:20

Received by: D.P.

Relinquished by: JO