


LB108414

7471B

INSTRUMENT ID: CV1

 μ Abs.:

55391



Linear

A= 0.0000e+000

B= 1.8176e-004

C= 8.8536e-003

Rho= 0.9998984

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	27	0.000	27				
0.2	0.200	0.214	0.014	1128	0.0 %	1128				
2.5	2.500	2.532	0.032	13879	0.0 %	13879				
5.0	5.000	4.939	-0.061	27122	0.0 %	27122				
7.5	7.500	7.425	-0.075	40804	0.0 %	40804				
10.0	10.000	10.077	0.077	55391	0.0 %	55391				

LB108414 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Mar 2020 11:31:16

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	27	-	0.00007471B	PPB		24 Mar 2020 11:37:53	S	Std
0.2 - 1	50.2	1128	-	0.20007471B	PPB		24 Mar 2020 11:40:10	S	Std
2.5 - 1	52.5	13879	-	2.50007471B	PPB		24 Mar 2020 11:42:27	S	Std
5.0 - 1	55	27122	-	5.00007471B	PPB		24 Mar 2020 11:44:44	S	Std
7.5 - 1	57.5	40804	-	7.50007471B	PPB		24 Mar 2020 11:47:00	S	Std
10.0 - 1	510	55391	-	10.00007471B	PPB		24 Mar 2020 11:49:17	S	Std
ICV73 - 1	ICV73	21059	3.8366	-7471B	PPB		24 Mar 2020 11:53:45	U	SMPL
ICB73 - 1	ICB73	-177	-0.0233	-7471B	PPB		24 Mar 2020 11:55:59	U	SMPL
CCV84 - 1	CCV84	27388	4.9869	-7471B	PPB		24 Mar 2020 11:58:15	U	SMPL
CCB84 - 1	CCB84	-217	-0.0306	-7471B	PPB		24 Mar 2020 12:00:30	U	SMPL
CRA - 1	CRA	1085	0.2061	-7471B	PPB		24 Mar 2020 12:02:46	U	SMPL
HighStd - 1	HighStd	55256	10.0523	-7471B	PPB		24 Mar 2020 12:05:01	U	SMPL
ChkStd - 1	ChkStd	37569	6.8374	-7471B	PPB		24 Mar 2020 12:07:16	U	SMPL
PB127771BL - 1	PBS	-359	-0.0564	-7471B	PPB		24 Mar 2020 12:09:34	U	SMPL
PB127771BS - 1	LCSS	18899	3.4440	-7471B	PPB		24 Mar 2020 12:11:52	U	SMPL
L2013-06 - 1	GHD1-1BC-IB-0-5	42541	7.7412	-7471B	PPB		24 Mar 2020 12:14:08	U	SMPL
L2013-07 - 1	GHD1-BC-IB-0-5	46938	8.5404	-7471B	PPB		24 Mar 2020 12:16:25	U	SMPL
L2013-08 - 1	GHD1-SB-01-0-52	19439	3.5421	-7471B	PPB		24 Mar 2020 12:18:44	U	SMPL
L2013-09 - 1	GHD1-SB-0203-0-11	33371	6.0744	-7471B	PPB		24 Mar 2020 12:21:02	U	SMPL
L2019-01 - 1	DRILL-CUTTINGS	224	0.0496	-7471B	PPB		24 Mar 2020 12:23:18	U	SMPL
CCV85 - 1	CCV85	27407	4.9904	-7471B	PPB		24 Mar 2020 12:25:35	U	SMPL
CCB85 - 1	CCB85	-135	-0.0157	-7471B	PPB		24 Mar 2020 12:27:50	U	SMPL
L2020-01 - 1	5TH-STREET-SUB	3447	0.6354	-7471B	PPB		24 Mar 2020 12:32:43	U	SMPL
L2020-01DUP - 1	5TH-STREET-SUBDUP	3407	0.6281	-7471B	PPB		24 Mar 2020 12:34:57	U	SMPL
L2020-01MS - 1	5TH-STREET-SUBMS	23498	4.2799	-7471B	PPB		24 Mar 2020 12:37:12	U	SMPL
L2020-01MSD - 1	5TH-STREET-SUBMSD	24936	4.5413	-7471B	PPB		24 Mar 2020 12:39:28	U	SMPL
L2020-01LX5 - 1		518	0.1030	-7471B	PPB		24 Mar 2020 12:41:44	U	SMPL
L2020-01A - 1		24507	4.4633	-7471B	PPB		24 Mar 2020 12:44:01	U	SMPL
CCV86 - 1	CCV86	27695	5.0427	-7471B	PPB		24 Mar 2020 12:46:17	U	SMPL
CCB86 - 1	CCB86	-129	-0.0146	-7471B	PPB		24 Mar 2020 12:48:34	U	SMPL