

LB108420


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37605



A= 0.0000e+000

B= 2.6732e-004

C= -5.4302e-003

Rho= 0.9999138

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	34	0.000	34				
0.2	0.200	0.196	-0.004	755	0.0 %	755				
2.5	2.500	2.502	0.002	9380	0.0 %	9380				
5.0	5.000	5.048	0.048	18905	0.0 %	18905				
7.5	7.500	7.403	-0.097	27712	0.0 %	27712				
10.0	10.000	10.047	0.047	37605	0.0 %	37605				

LB108420 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 24 Mar 2020 14:03:45							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type		
0.0 - 1	50	34	-	0.00007470A	PPB		24 Mar 2020 14:06:43	S	Std		
0.2 - 1	50.2	755	-	0.20007470A	PPB		24 Mar 2020 14:08:51	S	Std		
2.5 - 1	52.5	9380	-	2.50007470A	PPB		24 Mar 2020 14:10:59	S	Std		
5.0 - 1	55	18905	-	5.00007470A	PPB		24 Mar 2020 14:13:06	S	Std		
7.5 - 1	57.5	27712	-	7.50007470A	PPB		24 Mar 2020 14:15:13	S	Std		
10.0 - 1	510	37605	-	10.00007470A	PPB		24 Mar 2020 14:17:19	S	Std		
ICV74 - 1	ICV74	15006	4.0060	-7470A	PPB		24 Mar 2020 14:20:01	U	SMPL		
ICB74 - 1	ICB74	-171	-0.0511	-7470A	PPB		24 Mar 2020 14:22:07	U	SMPL		
CCV87 - 1	CCV87	18002	4.8069	-7470A	PPB		24 Mar 2020 14:24:13	U	SMPL		
CCB87 - 1	CCB87	-175	-0.0522	-7470A	PPB		24 Mar 2020 14:26:18	U	SMPL		
CRA - 1	CRA	670	0.1737	-7470A	PPB		24 Mar 2020 14:28:25	U	SMPL		
HighStd - 1	HighStd	38000	10.1527	-7470A	PPB		24 Mar 2020 14:30:31	U	SMPL		
ChkStd - 1	ChkStd	26073	6.9644	-7470A	PPB		24 Mar 2020 14:32:37	U	SMPL		
PB127775BL - 1	PBW	-273	-0.0784	-7470A	PPB		24 Mar 2020 14:34:45	U	SMPL		
PB127775BS - 1	LCSW	15485	4.1340	-7470A	PPB		24 Mar 2020 14:36:53	U	SMPL		
L2011-01 - 1	GHD1-WC-S-BC-IBC	-150	-0.0455	-7470A	PPB		24 Mar 2020 14:39:00	U	SMPL		
L2011-01DUP - 1	GHD1-WC-S-BC-IBCDUP	-25	-0.0121	-7470A	PPB		24 Mar 2020 14:41:08	U	SMPL		
L2011-01MS - 1	GHD1-WC-S-BC-IBCMS	13430	3.5847	-7470A	PPB		24 Mar 2020 14:43:15	U	SMPL		
L2011-01MSD - 1	GHD1-WC-S-BC-IBCMSD	13428	3.5841	-7470A	PPB		24 Mar 2020 14:45:22	U	SMPL		
L2011-02 - 1	GHD1-WC-S-SB-01	-162	-0.0487	-7470A	PPB		24 Mar 2020 14:47:29	U	SMPL		
CCV88 - 1	CCV88	18800	5.0202	-7470A	PPB		24 Mar 2020 14:49:35	U	SMPL		
CCB88 - 1	CCB88	-168	-0.0503	-7470A	PPB		24 Mar 2020 14:51:41	U	SMPL		
L2011-03 - 1	GHD1-WC-S-SB-0203	-34	-0.0145	-7470A	PPB		24 Mar 2020 14:53:47	U	SMPL		
L2020-04 - 1	5TH-STREET-SUB	-4	-0.0065	-7470A	PPB		24 Mar 2020 14:55:53	U	SMPL		
PB127776BL - 1	PBW	13	-0.0020	-7470A	PPB		24 Mar 2020 14:57:59	U	SMPL		
PB127776BS - 1	LCSW	15572	4.1573	-7470A	PPB		24 Mar 2020 15:00:06	U	SMPL		
L2021-07 - 1	MW-A	-162	-0.0487	-7470A	PPB		24 Mar 2020 15:02:12	U	SMPL		
L2021-07DUP - 1	MW-ADUP	-23	-0.0116	-7470A	PPB		24 Mar 2020 15:04:19	U	SMPL		
L2021-07MS - 1	MW-AMS	13326	3.5569	-7470A	PPB		24 Mar 2020 15:06:26	U	SMPL		
L2021-07MSD - 1	MW-AMSD	12998	3.4692	-7470A	PPB		24 Mar 2020 15:08:34	U	SMPL		
L2021-08 - 1	MW-B	-166	-0.0498	-7470A	PPB		24 Mar 2020 15:10:42	U	SMPL		
PB127777BL - 1	PBW	-11	-0.0084	-7470A	PPB		24 Mar 2020 15:12:48	U	SMPL		
CCV89 - 1	CCV89	18760	5.0095	-7470A	PPB		24 Mar 2020 15:14:54	U	SMPL		
CCB89 - 1	CCB89	-179	-0.0533	-7470A	PPB		24 Mar 2020 15:17:00	U	SMPL		
PB127777BS - 1	LCSW	13467	3.5946	-7470A	PPB		24 Mar 2020 15:19:08	U	SMPL		
L2011-04 - 1	GHD1-WC-C-01	310	0.0774	-7470A	PPB		24 Mar 2020 15:21:14	U	SMPL		
L2012-01 - 1	GHD1-GW-03-17	390	0.0988	-7470A	PPB		24 Mar 2020 15:23:20	U	SMPL		
L2019-02 - 1	WATER	-40	-0.0161	-7470A	PPB		24 Mar 2020 15:25:27	U	SMPL		
L2019-02DUP - 1	WATERDUP	-20	-0.0108	-7470A	PPB		24 Mar 2020 15:27:34	U	SMPL		
L2019-02MS - 1	WATERMS	14159	3.7796	-7470A	PPB		24 Mar 2020 15:29:41	U	SMPL		
L2019-02MSD - 1	WATERMSD	14654	3.9119	-7470A	PPB		24 Mar 2020 15:31:48	U	SMPL		
L2021-01 - 1	MW-A	25373	6.7773	-7470A	PPB		24 Mar 2020 15:33:56	U	SMPL		
L2021-04 - 1	MW-B	53993	14.4280	-7470A	PPB		24 Mar 2020 15:38:37	U	SMPL		
CCV90 - 1	CCV90	19164	5.1175	-7470A	PPB		24 Mar 2020 15:40:44	U	SMPL		
CCB90 - 1	CCB90	-23	-0.0116	-7470A	PPB		24 Mar 2020 15:42:52	U	SMPL		
L2011-01LX5 - 1		39	0.0050	-7470A	PPB		24 Mar 2020 15:44:58	U	SMPL		
L2011-01A - 1		15208	4.0600	-7470A	PPB		24 Mar 2020 15:47:03	U	SMPL		
L2021-07LX5 - 1		-20	-0.0108	-7470A	PPB		24 Mar 2020 15:49:10	U	SMPL		
L2021-07A - 1		15044	4.0161	-7470A	PPB		24 Mar 2020 15:51:16	U	SMPL		
L2019-02LX5 - 1		-63	-0.0223	-7470A	PPB		24 Mar 2020 15:53:22	U	SMPL		
L2019-02A - 1		15060	4.0204	-7470A	PPB		24 Mar 2020 15:55:28	U	SMPL		
PB127775TB - 1	PB127756TB	-74	-0.0252	-7470A	PPB		24 Mar 2020 15:57:35	U	SMPL		
PB127776TB - 1	PB127758TB	36	0.0042	-7470A	PPB		24 Mar 2020 15:59:43	U	SMPL		
L2021-04DLX2 - 1		26141	6.9826	-7470A	PPB		24 Mar 2020 16:01:50	U	SMPL		
CCV91 - 1	CCV91	19436	5.1902	-7470A	PPB		24 Mar 2020 16:03:58	U	SMPL		
CCB91 - 1	CCB91	50	0.0079	-7470A	PPB		24 Mar 2020 16:06:06	U	SMPL		