



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108442

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/25/2020	17:15
2	CAL2	1	Water	NA	NA	20.1	7.00	03/25/2020	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	03/25/2020	17:19
4	ICV	1	Water	NA	NA	20.1	6.99	03/25/2020	17:20
5	CCV1	1	Water	NA	NA	20.2	2.02	03/25/2020	18:15
6	L2031-01	1	Solid	20.03	20	20.5	8.15	03/25/2020	18:16
7	L2031-01DUP	1	Solid	20.04	20	20.6	8.16	03/25/2020	18:19
8	L2031-02	1	Solid	20.02	20	20.5	8.57	03/25/2020	18:20
9	L2031-03	1	Solid	20.03	20	20.2	9.34	03/25/2020	18:22
10	L2031-04	1	Solid	20.04	20	20.7	7.67	03/25/2020	18:25
11	L2031-05	1	Solid	20.04	20	21.0	9.11	03/25/2020	18:26
12	L2031-06	1	Solid	20.02	20	20.7	6.46	03/25/2020	18:29
13	L2031-12	1	Solid	20.03	20	20.5	8.98	03/25/2020	18:30
14	L2031-13	1	Solid	20.04	20	20.7	8.44	03/25/2020	18:32
15	L2031-14	1	Solid	20.02	20	20.5	6.04	03/25/2020	18:35
16	CCV2	1	Water	NA	NA	20.2	12.02	03/25/2020	18:36
17	L2031-15	1	Solid	20.02	20	20.9	8.71	03/25/2020	18:38
18	L2031-16	1	Solid	20.04	20	20.9	8.50	03/25/2020	18:40
19	L2031-17	1	Solid	20.02	20	20.7	8.81	03/25/2020	18:42
20	CCV3	1	Water	NA	NA	20.2	2.02	03/25/2020	18:45

18108442

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S L2030 WorkList ID : 137355 Department : Wet-Chemistry Date : 03-25-2020 15:56:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/25/2020	Solid	L2031-01	pH	Cool 4 deg C	PSEG01	F51	30899	03/25/2020	9045D
03/25/2020	Solid	L2031-02	pH	Cool 4 deg C	PSEG01	F51	30837	03/25/2020	9045D
03/25/2020	Solid	L2031-03	pH	Cool 4 deg C	PSEG01	F51	30926	03/25/2020	9045D
03/25/2020	Solid	L2031-04	pH	Cool 4 deg C	PSEG01	F51	30946	03/25/2020	9045D
03/25/2020	Solid	L2031-05	pH	Cool 4 deg C	PSEG01	F51	30944	03/25/2020	9045D
03/25/2020	Solid	L2031-06	pH	Cool 4 deg C	PSEG01	F51	30945	03/25/2020	9045D
03/25/2020	Solid	L2031-12	pH	Cool 4 deg C	PSEG01	F51	30942	03/25/2020	9045D
03/25/2020	Solid	L2031-13	pH	Cool 4 deg C	PSEG01	F51	30943	03/25/2020	9045D
03/25/2020	Solid	L2031-14	pH	Cool 4 deg C	PSEG01	F51	30908	03/25/2020	9045D
03/25/2020	Solid	L2031-15	pH	Cool 4 deg C	PSEG01	F51	30951	03/25/2020	9045D
03/25/2020	Solid	L2031-16	pH	Cool 4 deg C	PSEG01	F51	30952	03/25/2020	9045D
03/25/2020	Solid	L2031-17	pH	Cool 4 deg C	PSEG01	F51	30953	03/25/2020	9045D

Date/Time 03/25/2020 16:00

Received by: [Signature]

Relinquished by: [Signature]

Date/Time 03/25/2020 18:20

Received by: [Signature]

Relinquished by: [Signature]